



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED 8/1/14-CD  
 REPORT #7

**REVIEWED**  
 By Carol Day at 12:42 pm, Aug 27, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102472	PRINTER SN 08C.3556.272	DATE OF INSPECTION 07/21/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Greene County Sheriff's Office 1010 N. Boonville Springfield	TIME OF INSPECTION 8:56 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2219 SIMULATOR EXP DATE 04/23/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .100

TEST 2 ➔ .100

TEST 3 ➔ .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	3	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	3	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**


SIGNATURE *Ronald L. Killingsworth*

PRINT NAME  
 Ronald L. Killingsworth

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 230085 / 05/10/2015

TELEPHONE NUMBER  
 (417) 829-6216

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**RONALD L KILLINGSWORTH**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSA40 and 308.111 through 308.119 RSA60.

DATE 05/10/2013

NUMBER 230085

EXPIRES 05/10/2015

MO 880-071 (8-10)

*W. L. Killingsworth*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Shad Vookley*

Asst. Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LS-4 (8-10)



**GUTH LABORATORIES, INC.**

**1000 W. 10TH ST. • DENVER, CO 80202 • TEL: 303.733.1111**

### **CERTIFICATE OF ANALYSIS**

#### **Certified Alcohol Reference Solution for Simulator**

Random Samples of Lot Number 13200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President  
GUTH LABORATORIES, INC.**

***NIST Traceability:***

***Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.***

***All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.***

AS IU Serial no: 102472  
Version no: 532B

TEST RECORD 00081

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/21/14 08:56 .000  
Calibration:  
24 07/21/14 08:56 .100

Subject Name

Test 1

Subject I.D.

Killingsworth, Ronald L.

Operator Name, I.D.

230085 / 05-10-15

Location

GCSO

AS IU Serial no: 102472  
Version no: 532B

TEST RECORD 00082

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/21/14 08:57 .000  
Calibration:  
25 07/21/14 08:57 .100

Subject Name

Test 2

Subject I.D.

Killingsworth, Ronald L.

Operator Name, I.D.

230085 / 05-10-15

Location

GCSO

AS IU Serial no: 102472  
Version no: 532B

TEST RECORD 00083

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/21/14 08:59 .000  
Calibration:  
25 07/21/14 08:59 .100

Subject Name

Test 3

Subject I.D.

Killingsworth, Ronald L.

Operator Name, I.D.

230085 / 05-10-15

Location

GCSO

AS IV Serial no: 102472  
Version no: 537B

TEST RECORD 00084

Temp Date Time 210L

Air Blank:  
07/21/14 09:01 .000  
Calibration Check:  
25 07/21/14 09:01 .000

Subject Name

Blank Test

Subject I.D.

Killingsworth, Ronald L.

Operator Name, I.D.

230085 / 05-10-15

Location

GCSO

AS IV Serial no: 102472  
Version no: 537B

TEST RECORD 00086

Temp Date Time 210L

VOID: RFI  
12 07/21/14 09:02

Subject Name

RFI Test

Subject I.D.

Killingsworth, Ronald L.

Operator Name, I.D.

230085 / 05-10-15

Location

GCSO