MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 9:18 am, Jul 23, 2014

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
102473	08C.3527.124	07/21/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
Greene County Sheriff's Office 1010 N. Boonville Springfield	10:25 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Laboratories LOT # 13280 EXP. DATE 10/16/2015

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2219 SIMULATOR EXP DATE 04/23/2015

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .100	TEST 3 ☛ .100

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
1	1	2	3	1	4

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Ronald L. Killingsworth	Ronald L. Killingsworth
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
230085 / 05/10/2015	(417) 829-6216

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RONALD L KILLINGSWORTH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/10/2013

NUMBER 230085

EXPIRES 05/10/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (9-10)

LAB-4 (10-10)

GUTH LABORATORIES, INC.
[illegible] STREET • [illegible] PA [illegible] • TELEPHONE [illegible]

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Certillant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00088

Temp	Date	Time	g/210L
Air Blank:			
	07/21/14	10:25	.000
Calibration:			
23	07/21/14	10:25	.100

Subject Name
Test 1
Subject I.D.
Killingsworth, Ronald C.
Operator Name, I.D.
230085 / 05-10-15
Location
GCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00089

Temp	Date	Time	g/210L
Air Blank:			
	07/21/14	10:27	.000
Calibration:			
23	07/21/14	10:27	.100

Subject Name
Test 2
Subject I.D.
Killingsworth, Ronald C.
Operator Name, I.D.
230085 / 05-10-15
Location
GCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00090

Temp	Date	Time	g/210L
Air Blank:			
	07/21/14	10:28	.000
Calibration:			
23	07/21/14	10:28	.100

Subject Name
Test 3
Subject I.D.
Killingsworth, Ronald C.
Operator Name, I.D.
230085 / 05/10-15
Location
GCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00091

Temp	Date	Time	g/ 210L
Air Blank:			
	07/21/14	10:29	.000
Calibration Check:			
24	07/21/14	10:29	.000

Subject Name
Blank Test

Subject I.D.
Killingsworth, Ronald L.

Operator Name, I.D.
230085 / 05-10-15

Location
GCSO

AS IV Serial no: 102473
Version no: 532B

TEST RECORD 00092

Temp	Date	Time	g/ 210L
[illegible]			
2	07/[illegible]/14	10:[illegible]	

[illegible]
RFI Test

[illegible]
Killingsworth, Ronald L.

[illegible] I.D.
230085 / 05-10-15

[illegible]
GCSO