



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED REPORT 7
 By Carol Day at 2:47 pm, Dec 31, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 104658	PRINTER SN 096.3580.740	DATE OF INSPECTION 12/28/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 8920 Natural Bridge Bel-Ridge	TIME OF INSPECTION 4:13 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG306401 EXP. DATE 03/05/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .082

TEST 2 ➔ .081

TEST 3 ➔ .080

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Richard A Erwin
TYPE II PERMIT NUMBER/EXPIRATION DATE 204237/05-13-2016	TELEPHONE NUMBER (314) 889-2341

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 104658
Version no: 532B

TEST RECORD 00041
Temp Date Time 210L
s/

U010: BPI
12/28/14 16:17

Subject Name
Jon Doe
Subject I.D.
123456

Operator Name, I.D.
Erwin 284

Location
8920 Nknd

Druck

AS IV Serial no: 104658
Version no: 532B

AS IV Serial no: 104658
Version no: 532B

TEST RECORD 00039

Temp Date Time 210L
s/

Air Blank:
12/28/14 16:13 .000
Calibration Check:
21 12/28/14 16:13 .082

Subject Name

Jon Doe

Subject I.D.

123456

Operator Name, I.D.

Erwin 284

Location

8920 Nknd

Bridge

AS IV Serial no: 104658
Version no: 532B

TEST RECORD 00040

Temp Date Time 210L
s/

Air Blank:
12/28/14 16:16 .000
Calibration Check:
23 12/28/14 16:16 .080

Subject Name

Jon Doe

Subject I.D.

123456

Operator Name, I.D.

Erwin 284

Location

8920 Nknd

Bridge

AS IV Serial no: 104658
Version no: 532B

TEST RECORD 00039

Temp Date Time 210L
s/

Air Blank:
12/28/14 16:14 .000
Calibration Check:
22 12/28/14 16:14 .081

Subject Name

Jon Doe

Subject I.D.

123456

Operator Name, I.D.

Erwin 284

Location

8920 Nknd

Bridge



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 6-Mar-2013

Lot # AG306401

<u>Exp. Date</u>	<u>CYL. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
5-Mar-2015	108	Ethanol	0.082 ± 0.002 Br/AC (223 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Analyst:

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

RICHARD A ERWIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/13/2014


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 240237

David Vandenberg, acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/13/2016

MO 586-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **ERWIN, RICHARD**
 Permit No **240237**
 Date Issued **5/13/2014** Date Expires **5/13/2016**