



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 2:00 pm, May 19, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105446	PRINTER SN 099.3586.164	DATE OF INSPECTION 05/16/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 401 N Bynum, Lone Jack		TIME OF INSPECTION 1:15 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 13100 EXP. DATE 04/23/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN 094948 SIMULATOR EXP DATE 03/07/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .099	TEST 2 → .099	TEST 3 → .100
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Tim Cosner</i>	PRINT NAME Tim Cosner
TYPE II PERMIT NUMBER/EXPIRATION DATE 230113 06/12/2015	TELEPHONE NUMBER (816) 697-2417

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

800 NORTH 6TH STREET • HARRISBURG, PA 17111-4911 • TELEPHONE: 717-654-8670

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C \pm 0.2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Certified Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00151

Temp Date Time ^{g/} 210L

Blank:
05/16/14 13:15 .000
Calibration Check:
28 05/16/14 13:15 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Tim Casner

Location

401 N Bynum

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00152

Temp Date Time ^{g/} 210L

Air Blank:
05/16/14 13:17 .000
Calibration Check:
28 05/16/14 13:17 .099

Subject Name

Subject I.D.

Operator Name, I.D.

Tim Casner

Location

401 N Bynum

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00153

Temp Date Time ^{g/} 210L

Air Blank:
05/16/14 13:20 .000
Calibration Check:
29 05/16/14 13:20 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Tim Casner

Location

401 N Bynum

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00154

Temp Date Time ^{g/} 210L

VOID: RFI
12 05/16/14 13:22

Subject Name

Subject I.D.

Tim Casner

Operator Name, I.D.

401 N Bynum

Location



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

TIMOTHY J COSNER JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 5000, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013

NUMBER 230113

EXPIRES 06/12/2015

MO 580-0771 16-101

W. S. ...

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dani Webster

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RC-10)