

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 185447
Version no: 532B

TEST RECORD 00011
Temp Date Time s/
210L

Air Blank:
07/30/14 17:37 .000
Calibration Check:
28 07/30/14 17:37 .096

Subject Name
TEST RECORD 00009
Temp Date Time s/
210L

Air Blank:
07/30/14 17:34 .001
Calibration Check:
26 07/30/14 17:34 .098

Subject Name
Test 1
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Handyskill 378
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Test 3
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

AS IV Serial no: 185447
Version no: 532B

TEST RECORD 00010
Temp Date Time s/
210L

Air Blank:
07/30/14 17:35 .000
Calibration Check:
27 07/30/14 17:35 .057

Subject Name
TEST RECORD 00010
Temp Date Time s/
210L

Air Blank:
07/30/14 17:35 .000
Calibration Check:
27 07/30/14 17:35 .057

Subject Name
Test 2
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Handyskill 378
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Test 2
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

AS IV Serial no: 185447
Version no: 532B

TEST RECORD 00012
Temp Date Time s/
210L

VOID: NFI
12 07/30/14 17:40

Subject Name
RFI Test
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Handyskill 378
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
RFI Test
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
RFI Test
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

AS IV Serial no: 185447
Version no: 532B

TEST RECORD 00009
Temp Date Time s/
210L

Air Blank:
07/30/14 17:34 .001
Calibration Check:
26 07/30/14 17:34 .098

Subject Name
TEST RECORD 00009
Temp Date Time s/
210L

Air Blank:
07/30/14 17:34 .001
Calibration Check:
26 07/30/14 17:34 .098

Subject Name
Test 1
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Handyskill 378
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Subject Name
Test 3
Subject I.D. D30163
Operator Name, I.D. S-14-15

Location
5962 wall st
Operator Name, I.D. S-14-15

Operator Name

Location

Operator Name, I.D.

Subject I.D.

Subject Name

Calibration Check

Date

Time

Version no

Serial no



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

JOSEPH L HOUDYSHELL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2013

NUMBER 230163

EXPIRES 8/14/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HOUDYSHELL, JOSEPH
 Permit No 230163
 Date Issued 8/14/2013 Date Expires 8/14/2015