



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
 Instrument is repaired
 By Carol Day at 11:12 am, Jan 13, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and when instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105449	PRINTER SN 097.3585.012	DATE OF INSPECTION 01/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 200 Highlands Blvd. Dr. Manchester MO 63011		TIME OF INSPECTION 0230

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 19°C

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Wuth Labs LOT # 13210 EXP. DATE 7/29/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34°C SIMULATOR SN sch2293 SIMULATOR EXP DATE 11/14/14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .099% TEST 2 = .098% TEST 3 = .099%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Operates within 'DOH' specs, rules & regulations

INSPECTING OFFICER

SIGNATURE: P.O.D. Ebert

PRINT NAME: P.O.D. Ebert

TYPE II PERMIT NUMBER/EXPIRATION DATE: 230071 EXP 4/26/2015

TELEPHONE NUMBER: (636) 227-1410

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

500 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

DAVID J EBERT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/26/2013

NUMBER 230071

EXPIRES 04/26/2015

David Vesterberg

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RE-10)

AS IU Serial no: 105449
Version no: 532B

TEST RECORD 00104

Temp Date Time 210L

UOIR RTI
01/07/14 02:33

Subject Name

Subject I.D.

P.O. Blawie
Description Names, I.D.

Description

AS IU Serial no: 105449
Version no: 532B

TEST RECORD 00103

Temp Date Time 210L

Air Blank:
01/07/14 02:32 .000

Calibration Check:
22 01/07/14 02:32 .099

Subject Name

Subject I.D.

P.O. Blawie
Description Names, I.D.

Description

AS IU Serial no: 105449
Version no: 532B

TEST RECORD 00102

Temp Date Time 210L

Air Blank:
01/07/14 02:30 .000

Calibration Check:
22 01/07/14 02:30 .098

Subject Name

Subject I.D.

P.O. Blawie
Description Names, I.D.

Description

AS IU Serial no: 105449
Version no: 532B

TEST RECORD 00100

Temp Date Time 210L

Air Blank:
01/07/14 02:27 .000

Calibration Check:
21 01/07/14 02:27 .099

Subject Name

Subject I.D.

P.O. Blawie
Description Names, I.D.

Description