



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:31 am, Feb 10, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105449	PRINTER SN 097.3585.012	DATE OF INSPECTION 02/08/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 200 Highlands Blvd. Dr. Manchester MO 63011		TIME OF INSPECTION

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 22°C
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 13210 EXP. DATE 7/28/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 32°C SIMULATOR SN sd2293 SIMULATOR EXP DATE 11/14/2014

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = 0.100% TEST 2 = 0.101% TEST 3 = 0.101%

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Operated within "DOH" specs, rules and regulations

INSPECTING OFFICER	
SIGNATURE <u>P.O. Eben</u>	PRINT NAME <u>P.O. Eben</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>230071 EXP 4/26/2015</u>	TELEPHONE NUMBER <u>(636) 227-1410</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 185449
Version no: 532B

TEST RECORD 00107

Temp Date Time 210L 9/

Air Blank: 02/08/14 08:12 .000
Calibration Check: 22 02/08/14 08:12 .101

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 185449
Version no: 532B

TEST RECORD 00106

Temp Date Time 210L 9/

Air Blank: 02/08/14 08:11 .900
Calibration Check: 22 02/08/14 08:11 .100

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 185449
Version no: 532B

TEST RECORD 00109

Temp Date Time 210L 9/

VOID: RT1
12 02/08/14 08:15

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 185449
Version no: 532B

TEST RECORD 00108

Temp Date Time 210L 9/


Air Blank: 02/08/14 08:14 .000
Calibration Check: 22 02/08/14 08:14 .101

Subject Name


Subject I.D.

Operator Name: I.D.

Location



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
 TYPE II

DAVID J EBERT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV W/PRINTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/26/2013

NUMBER 230071

EXPIRES 04/26/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Ms W. Ebert
Ms W. Ebert

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



GUTH LABORATORIES, INC.

880 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13210** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.