



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RYAN A SHARP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2013

NUMBER 230233

EXPIRES 10/17/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SHARP, RYAN
Permit No 230233
Date Issued 10/17/2013 Date Expires 10/17/2015



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IU Serial no: 105452
Version no: 532B

TEST RECORD 00188

Temp Date Time 210L %/
Air Blank: 10/16/14 15:03 .000
Calibration Check: 21 10/16/14 15:03 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 105452
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00187

Temp Date Time 210L %/
Air Blank: 10/16/14 14:25 .000
Calibration Check: -20 10/16/14 14:25 .101

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 105452
Version no: 532B

TEST RECORD 00189

Temp Date Time 210L %/
Air Blank: 10/16/14 15:05 .000
Calibration Check: 22 10/16/14 15:05 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IU Serial no: 105452
Version no: 532B

TEST RECORD 00190

Temp Date Time 210L %/
VOID: KPI 12 10/16/14 15:06

Subject Name

Subject I.D.

Operator Name, I.D.

Location