



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Carol Day at 12:19 pm, Nov 24, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <b>107973</b>	PRINTER SN <b>099.3586.816</b>	DATE OF INSPECTION <b>11/19/2014</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>ST JOHN PD 8844 ST CHARLES RHD - ST JOHN MO 63114</b>		TIME OF INSPECTION <b>0500</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **25°C**

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY **0500 11/19/14**

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **CAULT LABORATORIES** LOT # **14030** EXP. DATE **1/20/16**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD3317** SIMULATOR EXP DATE **07/14/2015**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <b>.102</b>	TEST 2 • <b>.101</b>	TEST 3 • <b>.101</b>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>3</b>	(0-.04) <b>0</b>	(.05-.09) <b>1</b>	(.10-.14) <b>0</b>	(.15-.19) <b>1</b>	(OVER .19) <b>2</b>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

*Time changed due to Daylight Savings time*

**INSPECTING OFFICER**

SIGNATURE <b>SALVATORE MANUELE</b>	PRINT NAME <b>SALVATORE MANUELE</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>230095 05/28/2015</b>	TELEPHONE NUMBER <b>(314) 427-8700</b>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-504-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 167973  
Version no: 532B

TEST RECORD 00187

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/19/14 05:03 .000  
Calibration Check:  
25 11/19/14 05:03 .101

Subject Name

Simulator  
Subject I.D.

Operator Name, I.D.

Manuela 142

Location

8944 St Charles Rk Rd

STJOHN MO 63114

AS IV Serial no: 167973  
Version no: 532B

TEST RECORD 00186

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/19/14 05:01 .000  
Calibration Check:  
25 11/19/14 05:01 .101

Subject Name

Simulator  
Subject I.D.

Operator Name, I.D.

Manuela 142

Location

8944 St Charles Rk Rd

STJOHN MO 63114

AS IV Serial no: 167973  
Version no: 532B

TEST RECORD 00185

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
11/19/14 05:00 .000  
Calibration Check:  
25 11/19/14 05:00 .102

Subject Name

Simulator  
Subject I.D.

Operator Name, I.D.

Manuela 142

Location

8944 St Charles Rk Rd

STJOHN MO 63114

AS IV Serial no: 167973  
Version no: 532B

TEST RECORD 00188

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 11/19/14 05:04

Subject Name


Subject I.D.

Operator Name, I.D.

Manuela 142  
Location

8944 St Charles Rk Rd

STJOHN MO 63114



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**


**VALERIE MANUELE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013  
NUMBER 230095  
EXPIRES 05/28/2015

  
\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
  
Acting Director  
\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES