

AS IV Serial no: 107998
Version no: 532B

TEST RECORD 00265

Temp Date Time ^{s/} 210L

Air Blank:
10/12/14 08:59 .000
Calibration Check:
27 10/12/14 08:59 .095

Subject Name

TEST #2

Subject I.D.

Operator Name, I.D.

KESTERSON 279

Location

AS IV Serial no: 107998
Version no: 532B

TEST RECORD 00267

Temp Date Time ^{s/} 210L

VOID: RFI
12 10/12/14 09:02

Subject Name

RFI CHECK

Subject I.D.

Operator Name, I.D.

KESTERSON 279

Location

AS IV Serial no: 107998
Version no: 532B

TEST RECORD 00264

Temp Date Time ^{s/} 210L

Air Blank:
10/12/14 08:58 .000
Calibration Check:
26 10/12/14 08:58 .095

Subject Name

TEST #1

Subject I.D.

Operator Name, I.D.

KESTERSON 279

Location

AS IV Serial no: 107998
Version no: 532B

TEST RECORD 00266

Temp Date Time ^{s/} 210L

Air Blank:
10/12/14 09:01 .000
Calibration Check:
27 10/12/14 09:01 .095

Subject Name

TEST #3

Subject I.D.

Operator Name, I.D.

KESTERSON 279

Location

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

PAUL R KESTERSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/13/2014

NUMBER 240277

EXPIRES 6/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator KESTERSON, PAUL
Permit No 240277
Date Issued 6/13/2014 Date Expires 6/13/2016