



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
 By Carol Day at 2:05 pm, Jun 26, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107999	PRINTER SN 099.3586.796	DATE OF INSPECTION 6-12-14
LOCATION OF INSTRUMENT (STREET AND CITY) 211 W. WALNUT BOLIVAR MO 65613		TIME OF INSPECTION 1734

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER **GUTH LABORATORIES** LOT # **12170** EXP. DATE **9-5-14**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0** SIMULATOR SN **SD2749** SIMULATOR EXP DATE **4-23-15**

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .103	TEST 2 = .102	TEST 3 = .102
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Zach Palmer</i>	PRINT NAME ZACH PALMER
TYPE II PERMIT NUMBER/EXPIRATION DATE 240211 4-30-16	TELEPHONE NUMBER 417-326-5298

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 12170 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 11, 2012, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is September 5, 2014 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ZACH PALMER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/30/2014

NUMBER 240211

EXPIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PALMER, ZACH
Permit No 240211
Date Issued 4/30/2014 Date Expires 4/30/2016

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00024

Temp Date Time ^{s/} 210L

Air Blank:
06/12/14 17:37 .000
Calibration Check:
23 06/12/14 17:37 .102

Subject Name

MAINT.

Subject I.D.

Operator Name, I.D.

ZACH PALMER 240211

Location

211 W. WALNUT

BOLIVAR mo 65613

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00023

Temp Date Time ^{s/} 210L

Air Blank:
06/12/14 17:36 .000
Calibration Check:
22 06/12/14 17:36 .102

Subject Name

MAINT.

Subject I.D.

Operator Name, I.D.

ZACH PALMER 240211

Location

211 W. WALNUT

BOLIVAR mo 65613

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00022

Temp Date Time ^{s/} 210L

Air Blank:
06/12/14 17:34 .000
Calibration Check:
21 06/12/14 17:34 .103

Subject Name

MAINTENANCE

Subject I.D.

Operator Name, I.D.

ZACH PALMER 240211

Location

211 W. WALNUT

BOLIVAR mo 65613

AS IV Serial no: 187999
Version no: 532B

TEST RECORD 00025

Temp Date Time ^{s/} 210L

VOID: RFI
12 06/12/14 17:38

Subject Name

MAINT

Subject I.D.

Operator Name, I.D.

ZACH PALMER 240211

Location

211 W. WALNUT

Bolivar, mo 65613