



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 10:03 am, Oct 03, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108266	PRINTER SN 099.3586.822	DATE OF INSPECTION 10/02/2014
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 600 E. Walnut Columbia	TIME OF INSPECTION 3:20 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeter</u>	LOT # <u>AG414902</u> EXP. DATE <u>05/29/2016</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .098	TEST 2 ➔ .097	TEST 3 ➔ .097
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(.0-.04)	6	(.05-.09)	3	(.10-.14)	6	(.15-.19)	4	(OVER .19)	3
----------	---	----------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Scott Lenger
TYPE II PERMIT NUMBER/EXPIRATION DATE 240278 / 06/13/2016	TELEPHONE NUMBER (573) 874-7652

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



Airgas USA LLC (LAB)  
3500 Bernard Street  
St. Louis, Mo. 63103  
Ph: (314) 533-3100  
Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 2-Jun-2014

Lot # AG414902

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
29-May-2016	55	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2014.06.03 09:10:02 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Analyst: \_\_\_\_\_

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 108266  
Version no: 532B

TEST RECORD 00199

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/02/14 03:30 .000  
Calibration Check:  
25 10/02/14 03:30 .097

Subject Name

*TEST #3*

Subject I.D.

Operator Name, I.D.

*Scott Lewick #2040*

Location

*CPD*

AS IV Serial no: 108266  
Version no: 532B

TEST RECORD 00197

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/02/14 03:27 .000  
Calibration Check:  
25 10/02/14 03:27 .098

Subject Name

*TEST #1*

Subject I.D.

Operator Name, I.D.

*Scott Lewick #2090*

Location

*CPD*

AS IV Serial no: 108266  
Version no: 532B

TEST RECORD 00200

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 10/02/14 03:31

Subject Name

*RFI!*

Subject I.D.

Operator Name, I.D.

*Scott Lewick #2098*

Location

*CPD*

AS IV Serial no: 108266  
Version no: 532B

TEST RECORD 00198

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/02/14 03:28 .000  
Calibration Check:  
25 10/02/14 03:28 .097

Subject Name

*TEST #2*

Subject I.D.

Operator Name, I.D.

*Scott Lewick #209*

Location

*CPD*