



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:21 am, Jan 23, 2014 #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108388	St. Louis County Police Department	PRINTER SN 099-3586.815	DATE OF INSPECTION 01.16.2014
LOCATION OF INSTRUMENT (STREET AND CITY) MODOT T.M.C. : 14301 South Over 40 Road Town + Country, MO 63017			TIME OF INSPECTION 13:47

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) **18°C**
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER **GuH Laboratories** LOT # **13210** EXP. DATE **07.29.2015**
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) **34.0°C** SIMULATOR SN **SD269** SIMULATOR EXP DATE **07.16.2014**

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = .099	TEST 2 = .098	TEST 3 = .097
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	—	(0-.04)	—	(.05-.09)	—	(.10-.14)	—	(.15-.19)	—	(OVER .19)	—
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

New unit placed into service

INSPECTING OFFICER	
SIGNATURE P. D. Rose 1121	PRINT NAME Officer D. Rose, OSN 2721
TYPE II PERMIT NUMBER/EXPIRATION DATE 730253 / 11.12.2015	TELEPHONE NUMBER (314) 889.2341

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00008

Temp Date Time ^{s/} 210L

Air Blank:
01/16/14 13:47 .000
Calibration Check:
18 01/16/14 13:47 .099

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

ROSC 2721

Location

MODOT T.M.C.

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00009

Temp Date Time ^{s/} 210L

Air Blank:
01/16/14 13:49 .000
Calibration Check:
19 01/16/14 13:49 .098

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

ROSC 2721

Location

MODOT T.M.C.

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00010

Temp Date Time ^{s/} 210L

Air Blank:
01/16/14 13:50 .000
Calibration Check:
20 01/16/14 13:50 .097

Subject Name

N/A

Subject I.D.

N/A

Operator Name, I.D.

ROSC 2721

Location

MODOT T.M.C.

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00011

Temp Date Time ^{s/} 210L

Air Blank:
01/16/14 13:52 .000
Subject Test: Auto
22 01/16/14 13:52 .000

Subject Name

Self-Test

Subject I.D.

N/A

Operator Name, I.D.

ROSC 2721

Location

MODOT T.M.C.

AS IV Serial no: 108388
Version no: 532B

TEST RECORD 00012

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/16/14 13:53

Subject Name

RFI Test

Subject I.D.

N/A

Operator Name, I.D.

ROSC 2721

Location

MODOT T.M.C.



GUTH LABORATORIES, INC.

580 NORTH 6TH STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13210** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1216%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.


NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

DAVID M ROSE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/12/2013


NUMBER 230253

EXPIRES 11/12/2015

MO:580:0771 (6-10)



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY



,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (10-10)