



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 11:29 am, Sep 30, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------|----------------------------|----------------------------------|
| ALCO SENSOR IV SN 108389 | PRINTER SN 099.3586.804 | DATE OF INSPECTION 09/29/2014 |
|-----------------------------|----------------------------|----------------------------------|

| | |
|--|--------------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 303 E. 3rd Street Joplin | TIME OF INSPECTION 12:22 pm |
|--|--------------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc LOT # AG319103 EXP. DATE 03/10/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .099

TEST 2 ➔ .099

TEST 3 ➔ .098

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 1 | (0-.04) | 0 | (.05-.09) | 2 | (.10-.14) | 4 | (.15-.19) | 8 | (OVER .19) | 2 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
 Brett Davis

TYPE # PERMIT NUMBER EXPIRATION DATE
 230157/ 08-14-2015

TELEPHONE NUMBER
 (417) 623-3131

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
BRETT DAVIS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000, ALCO-SENSOR IV
 W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE: 8/14/2013
 NUMBER: 230157
 EXPIRES: 8/14/2015


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
Garl Vesterby
 acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (06-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DAVIS, BRETT
 Permit No 230157
 Date Issued 8/14/2013 Date Expires 8/14/2015



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 11-Jul-2013

Lot # AG319103

| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u> | <u>Certified Concentration</u> |
|------------------|------------------|---------------------|--------------------------------------|
| 10-Mar-2015 | 30 | Ethanol Nitrogen | 0.100 ± 2% BrAC (272 ppm) Balance |

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|-------------------|----------------------|-------------------|----------------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010595 | 208.9 ppm |
| EB0010561 | 103.7 ppm | EB0010562 | 104.9 ppm |
| EB0010681 | 52.22 ppm | EB0010579 | 52.94 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2013.07.11 10:33:53 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: 
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 108389
Version no: 532B

TEST RECORD 00192

Temp Date Time ^{g/} 210L

Air Blank:
09/29/14 12:22 .000
Calibration Check:
21 09/29/14 12:22 .099

Subject Name

Subject I.D.

Operator Name: I.D.

B B 9004

Location

Joplin Jail

AS IV Serial no: 108389
Version no: 532B

TEST RECORD 00193

Temp Date Time ^{g/} 210L

Air Blank:
09/29/14 12:24 .000
Calibration Check:
22 09/29/14 12:24 .099

Subject Name

Subject I.D.

Operator Name: I.D.

B B 9004

Location

Joplin Jail

AS IV Serial no: 108389
Version no: 532B

TEST RECORD 00194

Temp Date Time ^{g/} 210L

Air Blank:
09/29/14 12:26 .000
Calibration Check:
22 09/29/14 12:26 .098

Subject Name

Subject I.D.

Operator Name: I.D.

B B 9004

Location

Joplin Jail

AS IV Serial no: 108389
Version no: 532B

TEST RECORD 00195

Temp Date Time ^{g/} 210L

VOID: RFI
12 09/29/14 12:28

Subject Name

Subject I.D.

Operator Name: I.D.

B B 9004

Location

Joplin Jail