



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 1:01 pm, Jul 21, 2014

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108392	PRINTER SN 099.3586.814	DATE OF INSPECTION 07/14/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 5490 5th Street, Cottleville, Missouri 63304		TIME OF INSPECTION 6:41 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, INC LOT # 13280 EXP. DATE 10/16/2015

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34C SIMULATOR SN SD2588 SIMULATOR EXP DATE 01/29/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .095

TEST 2  .096

TEST 3  .096

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Nothing further.

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
 Jamie E. Snodgrass, 440

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 230051/03-21-2015

TELEPHONE NUMBER  
 (636) 498-6464

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

590 NORTH 6TH STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-684-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13280 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1217% (w/vol) ethyl alcohol. The expiration date for this lot number is October 16, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Certified Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

AS IV Serial no: 109392  
Version no: 532B

TEST RECORD 00042

Temp Date Time 210L

Air Blank:  
07/14/14 18:44 .000  
Calibration Check:  
27 07/14/14 18:44 .096

Subject Name  
Test 3  
Subject I.D.

Operator Name, I.D.  
J. Snodgrass, 440  
Location

Cottleville P.D.

AS IV Serial no: 109392  
Version no: 532B

TEST RECORD 00041

Temp Date Time 210L

Air Blank:  
07/14/14 18:42 .000  
Calibration Check:  
26 07/14/14 18:42 .096

Subject Name  
Test 2  
Subject I.D.

Operator Name, I.D.  
J. Snodgrass, 440  
Location

Cottleville P.D.

AS IV Serial no: 109392  
Version no: 532B

TEST RECORD 00040

Temp Date Time 210L

Air Blank:  
07/14/14 18:41 .000  
Calibration Check:  
25 07/14/14 18:41 .095

Subject Name  
Test 1  
Subject I.D.

Operator Name, I.D.  
J. Snodgrass, 440  
Location

Cottleville P.D.

AS IV Serial no: 109392  
Version no: 532B

TEST RECORD 00043

Temp Date Time 210L

VOID: RFI  
12 07/14/14 18:45

Subject Name  
RFI  
Subject I.D.

Operator Name, I.D.  
J. Snodgrass, 440  
Location

Cottleville P.D.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**JAMIE E SNODGRASS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 03/21/2013  
NUMBER 230051  
EXPIRES 03/21/2015

*W. A. Vesterberg*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*W. A. Vesterberg*  
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 582-0773 (8-10)

LAB-4 (RS-10)