MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

received 7/10/14-cd

INTOX EC/IR II MAINTENANCE REPORT

REPORT #3

REVIEWED
By Carol Day at 10:12 am, Jul 14, 2014

Complete this report at the time of the regular monthly preventive maintenance check (no [illegible] days). Complete this report whenever the instrument is serviced or repaired and whenever [illegible] into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN	NAME OF AGENCY	DATE OF INSPECTION
12677	University City Police	06/27/2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
6801 Delmar Blvd, University City		16:26 CDT

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

[X] DIAGNOSTIC RECORD

[X] BLANK CHECK	[X] CO2 CHECK
[X] FC 1 TEMP	[X] FLOW CHECK
[X] SRC TEMP	[X] FCB CHECK
[X] DET TEMP	[X] CRC COMP CHECK
[X] BT TEMP	[X] CRC CAL CHECK
[X] STD 2 TEMP	[X] PRINT TEST
[X] ETH CHECK	

BREATH ANALYZER ACCURACY STANDARDS

[ ] SIMULATOR SOLUTION [X] COMPRESSED ETHANOL-GAS MIXTURE

[X] STANDARD SUPPLIER INTOXIMETERS LOT# AG326803 EXP. DATE 09/25/2015

[ ] SIMULATOR TEMP (34°C ±0.2°C)	SIMULATOR S/N	SIMULATOR EXP DATE

[X] CALIBRATION CHECK - **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

[X] 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
[ ] 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
[ ] 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.100 g/210L	TEST 2 0.100 g/210L	TEST 3 0.100 g/210L

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS 2	0-.04 0	.05-.09 0	.10-.14 0	.15-.19 0	OVER .19 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE		PRINT FULL NAME
		COLP, DAVID
TIPP II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
230261	11/26/2015	(314) 725-2211

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, Missouri Department of Health and Senior Services,
Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901

Face This Side Down – This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
07/02/14

ARREST TIME: 23:00
SUBJECT NAME:
DOE/JOHN
DOB: 12/22/56 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
ASH/LANVE
OFFICER I.D.: 526
TESTING OFFICER:
ASH
OFFICER I.D.: 526
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
N
N

--- BREATH ANALYSIS ---

BLANK TEST	.000	00:15
INTERNAL STANDARD	VERIFIED	00:15
RADIO INTERFERENCE		

Operator Signature_____

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
07/01/14
23:31

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

rator Signature_____

2208-02

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
07/02/14

TESTING OFFICER:
ASH/LANCE
OFFICER I.D.: 526
PERMIT NUMBER: 24149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
CAL CHECK
N

--- SUPERVISOR MODE ---

BLANK TEST	.000	00:08
INTERNAL STANDARD	VERIFIED	00:08
EXTERNAL STANDARD	.103	00:08
BLANK TEST	.000	00:09
EXTERNAL STANDARD	.104	00:09
BLANK TEST	.000	00:10
EXTERNAL STANDARD	.103	00:11
BLANK TEST	.000	00:11

N = 3
SIM. = .1
AVG. = .1033

Operator Signature_____

2208