



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM

**RECEIVED**  
 By Carol Day at 11:29 am, Sep 30, 2014

**INTOX EC/IR II MAINTENANCE REPORT**

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|   |  |                                  |
|---|--|----------------------------------|
| INTOX EC/IR II SN<br>12688  | NAME OF AGENCY<br>Lee's Summit Police Dept | DATE OF INSPECTION<br>09/30/2014 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>10 NE Tudor Rd Lee's Summit |  | TIME OF INSPECTION<br>10:40 CDT  |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

|   |  |
|---|--|
| <input checked="" type="checkbox"/> BLANK CHECK | <input checked="" type="checkbox"/> CO2 CHECK      |
| <input checked="" type="checkbox"/> FC 1 TEMP   | <input checked="" type="checkbox"/> FLOW CHECK     |
| <input checked="" type="checkbox"/> SRC TEMP    | <input checked="" type="checkbox"/> FCB CHECK      |
| <input checked="" type="checkbox"/> DET TEMP    | <input checked="" type="checkbox"/> CRC COMP CHECK |
| <input checked="" type="checkbox"/> BT TEMP     | <input checked="" type="checkbox"/> CRC CAL CHECK  |
| <input checked="" type="checkbox"/> STD 2 TEMP  | <input checked="" type="checkbox"/> PRINT TEST     |
| <input checked="" type="checkbox"/> ETH CHECK   |  |

**BREATH ANALYZER ACCURACY STANDARDS**

|   |  |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION                           | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER    intoximeters | LOT# AG319902    EXP. DATE 07/18/2015                              |

|   |               |                    |
|---|---------------|--------------------|
| <input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C) | SIMULATOR S/N | SIMULATOR EXP DATE |
|---|---------------|--------------------|

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within  $\pm 5\%$  of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

|  |
|--|
| <input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE            |
| <input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE |
| <input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE            |

|                        |                        |                        |
|------------------------|------------------------|------------------------|
| TEST 1    0.079 g/210L | TEST 2    0.079 g/210L | TEST 3    0.079 g/210L |
|------------------------|------------------------|------------------------|

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

|          |   |       |   |         |   |         |   |         |   |          |   |
|----------|---|-------|---|---------|---|---------|---|---------|---|----------|---|
| REFUSALS | 0 | 0-.04 | 0 | .05-.09 | 0 | .10-.14 | 0 | .15-.19 | 0 | OVER .19 | 0 |
|----------|---|-------|---|---------|---|---------|---|---------|---|----------|---|

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

|                                 |                                      |
|---------------------------------|--------------------------------------|
| SIGNATURE<br>                   | PRINT FULL NAME<br>EASLEY, JOHN      |
| TYPE II PERMIT NUMBER<br>230184 | TELEPHONE NUMBER<br>( 816 ) 969-4223 |
| EXPIRATION DATE<br>09/04/2015   |                                      |

**RETURN COMPLETED REPORT TO THE:**

Breath Alcohol Program, Missouri Department of Health and Senior Services,  
 Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)  
 3600 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 18-Jul-2013

Lot # AG319902

| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u>    | <u>Certified Concentration</u>          |
|------------------|------------------|---------------------|---|
| 18-Jul-2015      | 108              | Ethanol<br>Nitrogen | 0.080 ± 0.002 BrAC (218 ppm)<br>Balance |

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|-------------------|----------------------|-------------------|----------------------|
| EB0010581         | 391.8 ppm            | EB0010603         | 392.5 ppm            |
| EB0010570         | 259.8 ppm            | EB0010559         | 258.9 ppm            |
| EB0010285         | 209.0 ppm            | EB0010595         | 208.9 ppm            |
| EB0010561         | 103.7 ppm            | EB0010562         | 104.9 ppm            |
| EB0010681         | 52.22 ppm            | EB0010679         | 52.94 ppm            |


**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2013.07.18 14:34:31 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Analyst: \_\_\_\_\_

*Rod Marsala*  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JOHN S EASLEY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):


**DATAMASTER, INTOXILYZER 5000, INTOX EC/IR II**


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/4/2013

NUMBER 230184

EXPIRES 9/4/2015

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
 acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (06-10)

**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator: EASLEY, JOHN  
 Permit No: 230184  
 Date Issued: 9/4/2013 Date Expires: 9/4/2015