



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**INTOX EC/IR II MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 1:34 pm, Dec 23, 2014

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12818	NAME OF AGENCY ST. ANN POLICE DEPT.	DATE OF INSPECTION 12/12/2014
----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 10405 Mo-180 St Ann	TIME OF INSPECTION 03:31 CST
---	---------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

- |   |  |
|---|--|
| <input checked="" type="checkbox"/> BLANK CHECK | <input checked="" type="checkbox"/> CO2 CHECK      |
| <input checked="" type="checkbox"/> FC 1 TEMP   | <input checked="" type="checkbox"/> FLOW CHECK     |
| <input checked="" type="checkbox"/> SRC TEMP    | <input checked="" type="checkbox"/> FCB CHECK      |
| <input checked="" type="checkbox"/> DET TEMP    | <input checked="" type="checkbox"/> CRC COMP CHECK |
| <input checked="" type="checkbox"/> BT TEMP     | <input checked="" type="checkbox"/> CRC CAL CHECK  |
| <input checked="" type="checkbox"/> STD 2 TEMP  | <input checked="" type="checkbox"/> PRINT TEST     |
| <input checked="" type="checkbox"/> ETH CHECK   |  |

**BREATH ANALYZER ACCURACY STANDARDS**

- |   |  |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION                 | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER intox | LOT# ag410601 EXP. DATE 04/16/2016                                 |
| <input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)       | SIMULATOR S/N SIMULATOR EXP DATE                                   |

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within +5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- |  |
|--|
| <input checked="" type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE |
| <input type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE            |
| <input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE            |

TEST 1 * 0.099 g/210L	TEST 2 * 0.099 g/210L	TEST 3 * 0.099 g/210L
-----------------------	-----------------------	-----------------------

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS	0	0-.04	1	.05-.09	2	.10-.14	0	.15-.19	1	OVER .19	0
----------	---	-------	---	---------	---	---------	---	---------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument tested within DOH specifications

**INSPECTING OFFICER**

SIGNATURE <i>William Goodrich</i>	PRINT FULL NAME Goodrich
TYPE II PERMIT NUMBER 240245	TELEPHONE NUMBER ( 314 ) 427-8000
EXPIRATION DATE 05/19/2016	


**RETURN COMPLETED REPORT TO THE:**

Breath Alcohol Program, Missouri Department of Health and Senior Services,  
 Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901



## Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:  
19 CSR 25-30.051 (4).



Technician Printed Name: Donald D. DeBoard

Technician Signature: Donald D. DeBoard

Date: 07/15/2014

Contact: Missouri Safety Center  
Breath-Alcohol Instrument Training Program  
660-543-4834



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 16-Apr-2014

Lot # AG410601

<b><u>Exp. Date</u></b> 16-Apr-2016	<b><u>Cyl. Type</u></b> 108	<b><u>Component</u></b> Ethanol Nitrogen	<b><u>Certified Concentration</u></b> 0.100 ± 2% BrAC (272 ppm) Balance
--	--------------------------------	--	---

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

<b><u>Serial No.</u></b>	<b><u>Concentration</u></b>	<b><u>Serial No.</u></b>	<b><u>Concentration</u></b>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2014.04.18 08:07:40 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Analyst:                     *Rod Marsala*                      
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**WILLIAM M GOODRICH**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and  
 and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/19/2014

NUMBER 240245

EXPIRES 5/19/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES