



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:15 pm, Nov 03, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 200059	NAME OF AGENCY Directorate of Emergency Services	DATE OF INSPECTION 11/01/2014
LOCATION OF INSTRUMENT (STREET AND CITY) Bldg 1000, Fort Leonard Wood		TIME OF INSPECTION 0719

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11/01/14 10719
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GUTH LABS INC LOT # 13010 EXP. DATE 01/09/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR5376 EXP. DATE 10/20/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 • .103	TEST 2 • .103	TEST 3 • .104
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 1	(.10-.14) 0	(.15-.19) 0	OVER .19 0
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

TESTED AND CERTIFIED AS WITHIN GUIDELINES ESTABLISHED BY DHSS.

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME ROBERT ISHMAEL
TYPE II PERMIT NUMBER/EXPIRATION DATE 230001 01/08/2015	TELEPHONE NUMBER (573) 596-1074

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

FORT LAUDERDALE
DIRECTORATE OF EMERGENCY SERVICES

BAC DATA MASTER SERIAL NUMBER 888888
11/21/14
07119

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM (04-07-2003): OKAY

REACTOR
SAMPLE CHAMBER: 43C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUANTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11/21/14 07:11:19
H1 11/21/14 07:11:19
P4 11/21/14 07:11:19

64-2

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

FORT LEONARD WOOD
DIRECTORATE OF EMERGENCY SERVICES

BAC DATAMASTER SERIAL NUMBER 200009
11/01/14

TESTING OFFICER:
ISHMAEL ROBERT
OFFICER I.D.: 18740
PERMIT NUMBER: 200001
EXPIRATION DATE: 01/09/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	07:22
INTERNAL STANDARD	VERIFIED	07:22
EXTERNAL STANDARD	.109	07:23
BLANK TEST	.000	07:23
EXTERNAL STANDARD	.109	07:24
BLANK TEST	.000	07:24
EXTERNAL STANDARD	.104	07:25
BLANK TEST	.000	07:25

N # 0
SIM # 1
AVD # 1000

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

FORT LEONARD WOOD
DIRECTORATE OF EMERGENCY SERVICES

BAC DATAMASTER SERIAL NUMBER 200009
11/01/14


ARREST TIME: 01:00
SUBJECT NAME:
NA
DOB: 01/01/01 SEX: M
STATE/D.L.: NA/NA
ARRESTING OFFICER:
NA

OFFICER I.D.: NA
TESTING OFFICER:
ISHMAEL ROBERT
OFFICER I.D.: 18740
PERMIT NUMBER: 200001
EXPIRATION DATE: 01/09/15
MISCELLANEOUS DATA:
RFI/TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	07:21
INTERNAL STANDARD	VERIFIED	07:21
RADIO INTERFERENCE		

Handwritten signature



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

ROBERT ALAN ISHMAEL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/08/2013

NUMBER 230001

EXPIRES 01/08/2015

MO 560-0771 (8-10)

W. W. S.

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Daryl W. Visher

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (P6-10)



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13010 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 14, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is January 9, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.