MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
By Carol Day at 10:21 am, Jul 22, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201147	Hermann Police Department	07/09/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
1902 Jefferson Street; Hermann	9:35 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 07-09-14 21:37
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 49 °C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 14110 EXP. DATE 05/01/2016
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2250 EXP. DATE 09/19/2014
- [ ] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.097	.097	.098

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	1	1	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ B. McGowan	Brian J. McGowan

TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
220254 09/11/2014	(573) 486-2211

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

State of Missouri
DEPARTMENT OF HEALTH

PERMIT
TYPE II

BRIAN MCGOWAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/11/2012

Number **220254**

Expires 09/11/2014

Director of State Public Health Laboratory

Director, Department of Health

MO 580-0771 (7-88)

Lab. 4 (R7-88)

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

HERMANN PD
BAC DATAMASTER SERIAL NUMBER 201147
07/09/14
21:37

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}→¯

Operator Signature B. McGowan 509

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

HERMANN PD
BAC DATAMASTER SERIAL NUMBER 201147
07/09/14

TESTING OFFICER:
MCGOWAN/B
OFFICER I.D.: 509
PERMIT NUMBER: 220254
EXPIRATION DATE: 09/11/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	21:56
INTERNAL STANDARD	VERIFIED	21:56
EXTERNAL STANDARD	.097	21:57
BLANK TEST	.000	21:57
EXTERNAL STANDARD	.098	21:58
BLANK TEST	.000	21:58
EXTERNAL STANDARD	.098	21:59
BLANK TEST	.000	21:59

N = 3
SIM. = .1
AVG. = .0976

Operator Signature B. McGowan 509

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

HERMANN PD
BAC DATAMASTER SERIAL NUMBER 201147
07/09/14

ARREST TIME: 21:00
SUBJECT NAME:
RFI/TEST
DOB: 01/01/75 SEX: M
STATE/D.L.: MO/12345
ARRESTING OFFICER:
N/A
OFFICER I.D.: 1
TESTING OFFICER:
MCGOWAN/B
OFFICER I.D.: 509
PERMIT NUMBER: 220254
EXPIRATION DATE: 09/11/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:02
INTERNAL STANDARD	VERIFIED	22:02
RADIO INTERFERENCE		

Operator Signature B. McGowan 509

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

HERMANN PD
BAC DATAMASTER SERIAL NUMBER 201147
07/09/14

ARREST TIME: 21:00
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/75 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
N/A
OFFICER I.D.: 1
TESTING OFFICER:
MCGOWAN/B
OFFICER I.D.: 509
PERMIT NUMBER: 220254
EXPIRATION DATE: 09/11/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:04
INTERNAL STANDARD	VERIFIED	22:04
SUBJECT SAMPLE	.000	22:04
BLANK TEST	.000	22:05

Operator Signature B. McGowan 509

2208-02