



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**ORIGINAL**

**RECEIVED**

By Carol Day at 12:53 pm, Sep 22, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201157	NAME OF AGENCY Cape Girardeau Police Department	DATE OF INSPECTION 09/13/2014
LOCATION OF INSTRUMENT (STREET AND CITY) BAT VAN (Checkpoint in Kennett, MO)		TIME OF INSPECTION 5:49 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09/13/2014 17:49
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repco	LOT # 14001 EXP. DATE 04/30/2016
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____ 34.0 _____ °C	SIMULATOR SN _____ SD2221 EXP. DATE 07/14/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .099	TEST 2 → .098	TEST 3 → .098
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Within DOH Specs

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>J. Lucas #1236</i>	PRINT FULL NAME Jeffrey D Lucas
TYPE II PERMIT NUMBER/EXPIRATION DATE 240223 05/10/2014	TELEPHONE NUMBER (573) 335-6621

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JEFFREY D LUCAS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/10/2014

NUMBER 240223

EXPIRES 5/10/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 589-0771 (6-10)

LAE-4 (P6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator: LUCAS, JEFFREY  
Permit No: 240223  
Date Issued 5/10/2014 Date Expires 5/10/2016

CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 14001**

**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 201157  
09/13/14

ARREST TIME: 10:00  
SUBJECT NAME:  
REF: TEST  
DOB: 01/01/81 SEX: M  
TATE: D.L.L.: HU/M123456789  
ARRESTING OFFICER:  
REF: TEST  
OFFICER I.D.: 000  
TESTING OFFICER:  
LUCAS, JEFFREY  
OFFICER I.D.: 226  
PERMIT NUMBER: 240223  
EXPIRATION DATE: 05/10/16  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	17:58
INTERNAL STANDARD	VERIFIED	17:58
RADIO INTERFERENCE		

Operator Signature

*J Lucas #226*

Printed on recycled paper with agri-based inks

CMSU 2208-96

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
BAC DATA MASTER SERIAL NUMBER 201157  
09/13/14

TESTING OFFICER:  
LUCAS, JEFFREY  
OFFICER I.D.: 226  
PERMIT NUMBER: 240223  
EXPIRATION DATE: 05/10/16  
MISCELLANEOUS DATA:  
PRECHECK  
POINT

--- SUPERVISOR NOISE ---

BLANK TEST	.000	17:57
INTERNAL STANDARD	VERIFIED	17:52
INTERNAL STANDARD	.099	17:52
BLANK TEST	.000	17:53
EXTERNAL STANDARD	.098	17:53
BLANK TEST	.000	17:54
EXTERNAL STANDARD	.096	17:55
BLANK TEST	.000	17:55

D - 0  
SPL - 1  
AVG - .0903

Operator Signature

*J Lucas #226*

Printed on recycled paper with agri-based inks

CMSU 2208-96

