MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 9:48 am, Sep 22, 2014 6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201171	Crestwood Police Department	09/17/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
1 Detjen Drive Crestwood	0:19 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 09/17/2014 12:19 am
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 48 °C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 14110 EXP. DATE 05/01/2016
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN DR5309 EXP. DATE 04/03/2015
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
  - [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
0.101	0.102	0.102

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Operating Properly

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Sgt K. Meier	Kent Meier
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
210102/01-03-2015	(314) 729-4800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CRESTWOOD POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 001171
09/17/14

TESTING OFFICER:
MEIER/KENT
OFFICER I.D.: 0254
PERMIT NUMBER: 21202
EXPIRATION DATE: 01/23/15
MISCELLANEOUS DATA:

---- SUPERVISOR MODE ----

BLANK TEST	.000	00:22
INTERNAL STANDARD	VERIFIED	00:22
EXTERNAL STANDARD	.101	00:23
BLANK TEST	.000	00:23
EXTERNAL STANDARD	.102	00:23
BLANK TEST	.000	00:24
EXTERNAL STANDARD	.102	00:25
BLANK TEST	.000	00:25

N = 3
SIM. = .1
AVG. = .1016

Operator Signature Sgt K. Meier 254

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CRESTWOOD POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 001171
09/17/14
00:25

---- DIAGNOSTIC CHECK ----

COMPUTER:	OKAY
PROGRAM (04-01-2000):	OKAY
HEATERS	
SAMPLE CHAMBER:	48C
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature Sgt K. Meier 254

2208-

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CRESTWOOD POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201171
09/17/14

ARREST TIME: 23:00
SUBJECT NAME:
TEST/RFI
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/1E3456
ARRESTING OFFICER:
MEIER/KENT
OFFICER I.D.: 0254
TESTING OFFICER:
MEIER/KENT
OFFICER I.D.: 0254
PERMIT NUMBER: 210102
EXPIRATION DATE: 01/00/15
MISCELLANEOUS DATA:
RFI TEST

---- BREATH ANALYSIS ----

BLANK TEST	.000	00:00
INTERNAL STANDARD	VERIFIED	00:00
RADIO INTERFERENCE		

Operator Signature Sgt K. Meier

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CRESTWOOD POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201171
09/17/14

ARREST TIME: 23:00
SUBJECT NAME:
TEST/BLANK
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/1E3456
ARRESTING OFFICER:
MEIER/KENT
OFFICER I.D.: 0254
TESTING OFFICER:
MEIER/KENT
OFFICER I.D.: 0254
PERMIT NUMBER: 210102
EXPIRATION DATE: 01/00/15
MISCELLANEOUS DATA:
BLANK TEST

---- BREATH ANALYSIS ----

BLANK TEST	.000	00:30
INTERNAL STANDARD	VERIFIED	00:00
SUBJECT SAMPLE	.000	00:31
BLANK TEST	.000	00:31

Operator Signature Sgt K. Meier

2208-

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 5, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **May 1, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FNI22211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

KENT MEIER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/03/2013

NUMBER 210102

EXPIRES 01/03/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)