



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 8:07 am, Feb 04, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201181	NAME OF AGENCY Nevada Police Department	DATE OF INSPECTION 01/29/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 120 S Ash St., Nevada		TIME OF INSPECTION 2:55 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/29/14 14:55
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc LOT # 13002 EXP. DATE 06/19/2014
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD22611 EXP. DATE 01/03/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ➡ .098	TEST 2 ➡ .098	TEST 3 ➡ .098
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <i>Det. Jens D Barclay</i>	PRINT FULL NAME Det. Jens D Barclay
TYPE II PERMIT NUMBER/EXPIRATION DATE 220229 09/07/2014	TELEPHONE NUMBER (417) 448-5152

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
NEVADA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201181
01/29/14
14:55

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS: OKAY
SAMPLE CHAMBER: 48c
FLOW DETECTOR: OKAY
PUMP: OKAY
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
1*8#%& /()@#%&^/*0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmnop
qrs tuvwxyz{|}~

Operator Signature *Det Paul Barclay*

2200-02

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
NEVADA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201181
01/29/14

TESTING OFFICER:
BARCLAY/JENS/D
OFFICER I.D.: 136
PERMIT NUMBER: 220029
EXPIRATION DATE: 09/07/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 14:57
INTERNAL STANDARD VERIFIED 14:57
EXTERNAL STANDARD .098 14:58
BLANK TEST .000 14:58
EXTERNAL STANDARD .098 14:59
BLANK TEST .000 14:59
EXTERNAL STANDARD .098 15:00
BLANK TEST .000 15:00

N = 3
SUM = .1
AVG. = .038

Operator Signature *Det Paul Barclay*

2200-02

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
NEVADA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201181
01/29/14

ARREST TIME: 14:30
SUBJECT NAME:
RFI/TEST
DOB: 01/29/14 SEX: M
STATE D.O.B.: MD/1
ARRESTING OFFICER:

OFFICER I.D.: 136
TESTING OFFICER:

OFFICER I.D.: 136
PERMIT NUMBER: 220029
EXPIRATION DATE: 09/07/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 15:00
INTERNAL STANDARD VERIFIED 15:00
SUBJECT SAMPLE .000 15:00
RADIO INTERFERENCE

Operator Signature *Det Paul Barclay*

2200-02

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



JENS D BARCLAY

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/07/2012

Number 220229

Expires 09/07/2014

Director of State Public Health Laboratory

Director, Department of Health