



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/4/14-CD REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

REVIEWED
By Carol Day at 9:56 am, Mar 14, 2014

DATAMASTER SN 201182	NAME OF AGENCY Hollister Police Department	DATE OF INSPECTION 02/25/2014
LOCATION OF INSTRUMENT (STREET AND CITY) #1 Gage Drive, Hollister		TIME OF INSPECTION 10:45 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 02/25/14, 2246 hrs.
- COMPUTER DETECTOR
- PROGRAM FILTERS
- HEATERS SAMPLE CHAMBER 49 °C QUARTZ STANDARD
- FLOW DETECTOR CALIBRATION
- PUMP HIGH SPEED PRINTER

- INDICATOR LIGHTS
- SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 13002 EXP. DATE 06/19/2015
- SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2732 EXP. DATE 08/23/2014

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	.099%	TEST 2	.099%	TEST 3	.099%
--------	-------	--------	-------	--------	-------

- PERFORM R.F.I. TEST (PRINTOUT ATTACHED)
- INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(.0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	2	OVER .19	1
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
Instrument Operating within DOHSS Specs and guidelines

INSPECTING OFFICER	
SIGNATURE <i>Timothy E. Matthews</i>	PRINT FULL NAME Sgt. Timothy E. Matthews
TYPE II PERMIT NUMBER/EXPIRATION DATE 220253 09/11/2014	TELEPHONE NUMBER (417) 334-6565

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919.876-5490

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013

The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN
HOLLISTER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 28168
02/25/14

ARREST TIME: 22:00

SUBJECT NAME:

TEST

DOB: 01/21/81 SEX: M

STATE/C.L.: MO/12345

ARRESTING OFFICER:

TEST

OFFICER I.D.: 4

TESTING OFFICER:

MATTHEWS/T/E

OFFICER I.D.: 562

PERMIT NUMBER: 288259

EXPIRATION DATE: 09/11/14

MISCELLANEOUS DATA:

RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST

DOB

2318

INTERNAL STANDARD

VERIFIED

2318

RADIO INTERFERENCE

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
HOLLISTER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 22118A
02/25/14
221186

--- SERVICITY CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2007): OKAY

HEATERS
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`ab cde fgh i j k l m n o
p q r s t u v w x y z { | } ~ "

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
HOLLISTER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 22118A
02/25/14

TESTING OFFICER:
MATTHEWS/JME
OFFICER I.D.# 582
PERMIT NUMBER: 220255
EXPIRATION DATE: 09/11/14
MISCELLANEOUS DATA

--- SUPERVISOR MODE ---

BLANK TEST	.822	22:32
INTERNAL STANDARD	VERIFIED	22:32
EXTERNAL STANDARD	.835	22:33
BLANK TEST	.828	22:33
EXTERNAL STANDARD	.839	22:34
BLANK TEST	.802	22:34
EXTERNAL STANDARD	.839	22:35
BLANK TEST	.806	22:35

N = 3
SIM. = .1
AVG. = .839

Operator Signature



State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



TIMOTHY E MATTHEWS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/11/2012

Number 220253

Expires 09/11/2014

MO 589-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (87-88)