



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 11:08 am, Nov 13, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201182	NAME OF AGENCY Hollister Police Department	DATE OF INSPECTION 11/10/2014
LOCATION OF INSTRUMENT (STREET AND CITY) #1 Gage Drive, Hollister		TIME OF INSPECTION 11:14 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11/10/2014 , 2314
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc. LOT # 14001 EXP. DATE 04/30/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2732 EXP. DATE 10/06/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .098%	TEST 2 <input checked="" type="checkbox"/> .098%	TEST 3 <input checked="" type="checkbox"/> .098%
--	--	--

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
 (USE OTHER SIDE IF NECESSARY).

Instrument Operating within DOHSS Specs and guidelines

INSPECTING OFFICER

SIGNATURE <i>Timothy E. Matthews</i>	PRINT FULL NAME Sgt. Timothy E. Matthews
PERMIT NUMBER/EXPIRATION DATE 240373 10/17/2016	TELEPHONE NUMBER (417) 334-6565

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN
DEPARTMENT OF LABOR RELATIONS

BAC DATA MASTER SERIAL NUMBER 00-100
11/10/14

TESTING OFFICER:
MATTHEWS

OFFICER I.D. NO. V

PERMIT NUMBER: 000070

EXPIRATION DATE: 12/17/14

MISCELLANEOUS NOTES

--- SUPERVISOR NOTE ---

BLANK TEST	.000	20:08
INTERNAL STANDARD	VERIFIED	20:08
EXTERNAL STANDARD	.000	20:09
BLANK TEST	.000	20:09
EXTERNAL STANDARD	.000	20:10
BLANK TEST	.000	20:10
EXTERNAL STANDARD	.000	20:11
BLANK TEST	.000	20:11

N = 3
S.D. = .1
AVG. = .000

Operator Signature

2208-4

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN
DEPARTMENT OF LABOR RELATIONS

BAC DATA MASTER SERIAL NUMBER 00-100
11/10/14

20:14

--- SYSTEMATIC CHECK ---

COMPUTER: OKAY

PROGRAM (00-07-00001): OKAY

WEETERS

PARAMETER CHANGES: OKAY

FLAME DETECTOR: OKAY

PUMP

ATOM SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&*()-_~|}{~<=>?[]\;:'"~
HIJKLMNOPQRSTUVWXYZ 1234567890~
~
~
~

Operator Signature

2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
HOLLISTER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER P01182
11/16/14

ARREST TIME: 23:09

SUBJECT NAME:

TEST

DOB: 01/01/01 SEX: M

STATE/REG. NO: MO/12345

ARRESTING OFFICER:

TEST

OFFICER I.D. #

TESTING OFFICER:

MATTHEWS/T/E

OFFICER I.D. # 502

PERMIT NUMBER: P49373

EXPIRATION DATE: 10/17/16

MISCELLANEOUS DATA:

RFI TEST

----- BREATH ANALYSIS -----

BLANK TEST	.000	23:42
INTERNAL STANDARD	VERIFIED	23:49
RADIO INTERFERENCE		

Operator Signature





STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

TIMOTHY E MATTHEWS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2014

NUMBER 240373

EXPIRES 10/17/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 589-0771 (3-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **MATTHEWS, TIMOTHY**
Permit No **240373**
Date Issued **10/17/2014** Date Expires **10/17/2016**



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations:
19 CSR 25-30.051 (4).

Checked: 10/06/2014 Expires: 10/06/2015
MSC Tech: DRL
Temp: 34.00 Digital Therm. SN: 093752
Agency: Hollister Police Dept.



SD 2732



Technician Printed Name: DAN LUCAS

Technician Signature: [Signature]

Date: 10/6/14

Contact: Missouri Safety Center
Breath-Alcohol Instrument Training Program
660-543-4834