



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED 3/14/14-CD REPORT #6

REVIEWED
 By Carol Day at 3:45 pm, Mar 31, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201183	NAME OF AGENCY State of Missouri Rogersville P.D.	DATE OF INSPECTION 03/08/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 211 E. Center St, Rogersville, MO 65742		TIME OF INSPECTION 11:09 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 2309 hours 03/08/14
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories Inc. LOT # 13100 EXP. DATE 04/23/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN DR4330 EXP. DATE 09/19/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → .095%	TEST 2 → .097%	TEST 3 → .097%
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Brad Obregon
TYPE II PERMIT NUMBER/EXPIRATION DATE 220166 07/06/2014	TELEPHONE NUMBER (417) 753-2884
RETURN COMPLETED REPORT TO THE:	Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

600 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER CALIBR
#0006104
28115

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2000): OKAY

HEATERS

SAMPLE CHAMBER: OK

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z 0 1 2 3 4 5 6 7 8 9

Operator Signature

2208-02

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER CALIBR
00108104

TESTING OFFICER:
DREGORY BRADY
OFFICER T. L. P. 751
PERMIT NUMBER: 220166
EXPIRATION DATE: 07-26-14
MISCELLANEOUS DATA:

--- SUPERVISOR PAGE ---

BLANK TEST	1.000	20120
INTERNAL STANDARD	VERIFIED	20120
EXTERNAL STANDARD	1.000	20120
BLANK TEST	1.000	20120
EXTERNAL STANDARD	1.000	20120
BLANK TEST	1.000	20120
EXTERNAL STANDARD	1.000	20120
BLANK TEST	1.000	20120

R = 0
SIM. = .1
AVG. = .9960

Operator Signature

BAC DataMaster Evidence Ticket


STATE OF MISSISSIPPI

STATE DEPARTMENT OF TRANSPORTATION
3000 N. GULF BLVD
MOBILE, AL 36688

ARREST NUMBER: 0000000000
SUBJECT NAME: [REDACTED]
ARREST DATE: 08/08/2014
ARREST TIME: 14:25
ARRESTING OFFICER: [REDACTED]
OFFICER ID: [REDACTED]
OFFICER NAME: [REDACTED]
OFFICER TITLE: [REDACTED]
OFFICER AGENCY: [REDACTED]
OFFICER PHONE: [REDACTED]
OFFICER FAX: [REDACTED]
OFFICER EMAIL: [REDACTED]
OFFICER ADDRESS: [REDACTED]
OFFICER CITY: [REDACTED]
OFFICER STATE: [REDACTED]
OFFICER ZIP: [REDACTED]

--- DATA ---

08/08/2014	14:25	0000000000
08/08/2014	14:25	0000000000
08/08/2014	14:25	0000000000

Operator Signature 

2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRAD OBREGON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 07/06/2012

Number 220166

Expires 07/06/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)