



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 201193
05/19/14
81781

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 40c

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstu
vwxyz{|}~"

Operator Signature



2209-02

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 201193
05/19/14


TESTING OFFICER:
CORRECTION AREA #
OFFICER I.D. # 751
PERMIT NUMBER: 220166
EXPIRATION DATE: 07/06/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	000	01:14
INTERNAL STANDARD	001	01:14
EXTERNAL STANDARD	002	01:15
BLANK TEST	003	01:17
EXTERNAL STANDARD	004	01:16
BLANK TEST	005	01:17
EXTERNAL STANDARD	006	01:18
BLANK TEST	007	01:18

= 3
SEK. # 11
AVG. # 1.3566

Operator Signature



Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 201183
05-18-11

ARREST TIME: 00:00
SUBJECT NAME:
RFL/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO-12345
ARRESTING OFFICER:
CORECON/BRAD/A
OFFICER I.D.: 751
TESTING OFFICER:
CORECON/BRAD/A
OFFICER I.D.: 751
PERMIT NUMBER: 220166
EXPIRATION DATE: 07/06/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	1.000	01:27
INTERNAL STANDARD	VERIFIED	01:27
RADIO INTERFERENCE		

Operator Signature 
2200-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRAD OBREGON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 07/06/2012

Number 220166

Expires 07/06/2014

Director of State Public Health Laboratory

Director, Department of Health

MO 68D-0771 (7-88)

Lab. 4 (R7-68)