



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
 DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED  
 By Carol Day at 8:32 am, Jun 16, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>201186</b>	NAME OF AGENCY <b>ROLLA POLICE DEPT.</b>	DATE OF INSPECTION <b>06-09-14</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>1007 N. ELM ST ROLLA, MO 65401</b>		TIME OF INSPECTION <b>0433 HRS.</b>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>06-09-14 0433 HRS</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **REPCO MARKETING, INC.** LOT # **13001** EXP. DATE **03-07-15**

SIMULATOR TEMP (34°C ± 0.2°C) **34** °C SIMULATOR SN **SD2725** EXP. DATE **07-10-14**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← <b>.095</b>	TEST 2 ← <b>.096</b>	TEST 3 ← <b>.096</b>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<b>1</b> (0-.04)	<b>0</b> (.05-.09)	<b>1</b> (.10-.14)	<b>2</b> (.15-.19)	<b>0</b> OVER .19	<b>0</b>
----------	------------------	--------------------	--------------------	--------------------	-------------------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE <b>Kenneth L. Moberly</b>	PRINT FULL NAME <b>KENNETH L. MOBERLY</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240135 / 04-03-16</b>	TELEPHONE NUMBER <b>(573) 308-1213</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13001**

**EXPIRATION DATE: March 7, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013  
The expiration date for this lot number is March 7, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

IN THE CIRCUIT COURT OF PHELPS COUNTY, MISSOURI  
ASSOCIATE DIVISION

STATE OF MISSOURI, Plaintiff )  
vs. ) Case No. \_\_\_\_\_  
\_\_\_\_\_, Defendant )

AFFIDAVIT PURSUANT TO SECTION 490.692 RSMo

Before me, the undersigned authority, personally appeared KENNETH L. MOBERLY, who, being by me duly sworn, deposes as follows: My name is KENNETH L. MOBERLY. I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am a "Type II Breathalyzer Operator," authorized by the Missouri Department of Health to carry out maintenance checks on a breathalyzer, and I am the custodian of the records of the breathalyzer for the Rolla Police Department, Rolla, Missouri. Attached hereto are \_\_\_\_\_ pages of records kept by the Rolla Police Department, Rolla, Missouri, in the regular course of business of the Rolla Police Department, Rolla, Missouri, for an employee or representative of the Rolla Police Department, Rolla, Missouri, with knowledge of the act, event, condition, opinion, or diagnosis recorded to make the records or to transmit information thereof to be included in such record; and the records was made at or near the time of the act, event, condition, opinion, or diagnosis. The records attached hereto are the original or exact duplicates of the original.

Kenneth L. Moberly  
Affiant

In witness whereof I have hereunto subscribed my name and affixed my official seal this 12<sup>th</sup> day of June, 2014.

Charles T. Satterfield  
Notary

My Commission Expires:

1-22-2018



CHARLES T. SATTERFIELD  
My Commission Expires  
January 22, 2018  
Dent County  
Commission #14425039

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186  
06/09/14

TESTING OFFICER:  
MOBERLY/KENNETH/L  
OFFICER I.D.: 998  
PERMIT NUMBER: 240135  
EXPIRATION DATE: 04/03/16  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE

--- SUPERVISOR MODE ---

BLANK TEST	.000	04:36
INTERNAL STANDARD	VERIFIED	04:36
EXTERNAL STANDARD	.095	04:36
BLANK TEST	.000	04:37
EXTERNAL STANDARD	.096	04:38
BLANK TEST	.000	04:39
EXTERNAL STANDARD	.096	04:39
BLANK TEST	.000	04:40

N = 3  
SIM. = .1  
AVG. = .0956

Operator Signature

*Kenneth L. Moberly*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186  
06/09/14

ARREST TIME: 03:00  
SUBJECT NAME:  
MOBERLY/KENNETH/L  
DOB: 11/26/63 SEX: M  
STATE/D.L.: MO/123456789  
ARRESTING OFFICER:  
MOBERLY/KENNETH/L  
OFFICER I.D.: 998  
TESTING OFFICER:  
MOBERLY/KENNETH/L  
OFFICER I.D.: 998  
PERMIT NUMBER: 240135  
EXPIRATION DATE: 04/03/16  
MISCELLANEOUS DATA:  
MONTHLY RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	04:43
INTERNAL STANDARD	VERIFIED	04:43
RADIO INTERFERENCE		

Operator Signature

*Kenneth L. Moberly*

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
ROLLA POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201186

06/09/14

04:33

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS  
SAMPLE CHAMBER: 49c

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstu  
vwxyz{|}~

Operator Signature

*Kenneth L. Moberly*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**KENNETH L MOBERLY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014

NUMBER 240135

EXPIRES 4/3/2016

MO 580-0771 (8-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R8-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **MOBERLY, KENNETH**  
Permit No **240135**  
Date issued **4/3/2014**    Date Expires **4/3/2016**