



RECEIVED
By Carol Day at 11:56 am, Apr 08, 2014
Missouri Public Health Laboratory



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201188	NAME OF AGENCY Kearney Police Department	DATE OF INSPECTION 04-08-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 725 WEST 92 HIGHWAY, Kearney, Mo, 64060		TIME OF INSPECTION 1010
CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.		
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 04-08-2014 @ 1010
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD	
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES, INC. LOT # 13100		EXP. DATE 04-23-2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C		SIMULATOR SN SD2751 EXP. DATE 12-17-2014
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)		
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)		
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE		
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE		
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		
TEST 1 ← .098	TEST 2 ← .098	TEST 3 ← .098
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)		
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)		
REFUSALS 5 (0-.04) 0	(.05-.09) 0	(.10-.14) 1
(.15-.19) 0	OVER .19 1	
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).		
INSPECTING OFFICER		
SIGNATURE Derick Hager #981	PRINT FULL NAME DERICK E. HAGER	
TYPE II PERMIT NUMBER/EXPIRATION DATE 230186 / 09-04-2015	TELEPHONE NUMBER (816) 903-4768	
RETURN COMPLETED REPORT TO THE:	Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901	



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201188
04/08/14

ARREST TIME: 09:50
SUBJECT NAME:
RFICHECK
DOB: 10/18/81 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
HAGER/DERICK/E
OFFICER I.D.: 0981
TESTING OFFICER:
HAGER/DERICK/E
OFFICER I.D.: 0981
PERMIT NUMBER: 230186
EXPIRATION DATE: 09/04/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 10:13
INTERNAL STANDARD VERIFIED 10:13
RADIO INTERFERENCE

OPERATOR SIGNATURE

Devil dog #981

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201188
04/08/14
10:10

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 49c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

OPERATOR SIGNATURE

Devil dog #981

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201188
04/08/14

ARREST TIME: 10:00
SUBJECT NAME:
HAGER/DERICK/E
DOB: 10/19/81 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
HAGER/DERICK/E
OFFICER I.D.: 0981
TESTING OFFICER:
HAGER/DERICK/E
OFFICER I.D.: 0981
PERMIT NUMBER: 230186
EXPIRATION DATE: 09/04/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:22
INTERNAL STANDARD	VERIFIED	10:22
SUBJECT SAMPLE	.000	10:23
BLANK TEST	.000	10:24

OPERATOR SIGNATURE

Devil Hage #981

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201188
04/08/14

TESTING OFFICER:
HAGER/DERICK/E
OFFICER I.D.: 0981
PERMIT NUMBER: 230186
EXPIRATION DATE: 09/04/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:16
INTERNAL STANDARD	VERIFIED	10:16
EXTERNAL STANDARD	.098	10:16
BLANK TEST	.000	10:17
EXTERNAL STANDARD	.098	10:17
BLANK TEST	.000	10:18
EXTERNAL STANDARD	.098	10:18
BLANK TEST	.000	10:19

N = 3
SIM. = .1
AVG. = .098

OPERATOR SIGNATURE

Devil Hage #981

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DERICK E HAGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/4/2013

NUMBER 230186

EXPIRES 9/4/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HAGER, DERICK
 Permit No 230186
 Date Issued 9/4/2013 Date Expires 9/4/2015