



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 9:31 am, May 27, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201196	NAME OF AGENCY Richmond Police Department	DATE OF INSPECTION 05/22/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 207 Summit Street, Richmond, MO 64085		TIME OF INSPECTION 5:50 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 05/22/2014 1750 hours
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repro Marketing Inc. LOT # 13001 EXP. DATE 03/07/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD 2670 EXP. DATE 01/17/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> .097	TEST 2 <input checked="" type="checkbox"/> .097	TEST 3 <input checked="" type="checkbox"/> .098
---	---	---

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operates within the Department of health & Senior Services standards.

INSPECTING OFFICER

SIGNATURE <i>Christopher S. Bagley</i>	PRINT FULL NAME Christopher Scott Bagley
TYPE II PERMIT NUMBER/EXPIRATION DATE 220360 / 10/12/2014	TELEPHONE NUMBER (816) 776-3575

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-168 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13001
EXPIRATION DATE: March 7, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1215 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 8, 2013
The expiration date for this lot number is March 7, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
RICHMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER: 226196
05/20/14

YES/NO OFFICER:
DAILY/DRY STOPPER/S
OFFICER: J. B. S. 196
PERMIT NUMBER: 221360
EXPIRATION DATE: 10-12/14
MISCELLANEOUS DATA:

--- SUPERVISOR NAME ---

ALCOHOL TEST	1.000	17:50
HYDRONAL STANDARD	VERIFY	17:50
EXTERNAL STANDARD	1.057	17:54
BLANK 118	1.000	17:54
INTERNAL STANDARD	1.097	17:55
BLANK 121	1.000	17:55
INTERNAL STANDARD	1.000	17:56
BLANK 124	1.000	17:56

4 = 0
50% = 1.1
90% = 1.0973

Signature Christopher S. Bayle

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
RICHMOND POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER: 226196
05/20/14
17:58

--- SUPERVISOR NAME ---

COMPUTER:	OKAY
PROGRAM (04-07-00001)	OKAY
MEMBERS	
PUMP: 04/05/04	450
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
TEMPERATURE:	OKAY
STORAGE:	OKAY
MOBILE STANDARD:	OKAY
CELL BATTERY:	OKAY

QUALITY TEST

THIS IS THE ONLY QUALITY TEST THAT IS REQUIRED TO BE PERFORMED ON ALL BAC DATAMASTER DEVICES. IT IS THE RESPONSIBILITY OF THE OPERATOR TO PERFORM THIS TEST.

Operator Signature Christopher S. Bayle

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
RICHMOND POLICE DEPT (BLND)

OFFICER POSITION SERIAL NUMBER 26196
05/22/14

ARREST NUMBER 11126
SUBJECT NAME
DOB
DOB RANGE
STATE/DIVISION NUMBER
PROPERTY IDENTIFICATION
VEHICLE/OWNER INFORMATION
OFFICER I.D. # 106
TESTING OFFICER
LABILEY/OWNER INFORMATION
OFFICER I.D. # 106
PERMIT NUMBER 201006
INSTRUMENT SERIAL # 101210
MISCELLANEOUS DATA

----- BREATH ANALYSIS -----

BLANK YES
INTERNAL STANDARD 10:06
RADIO INTERFERED VERIFIED 10:06

Operator Signature Christopher S. Bagley

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



CHRISTOPHER S BAGLEY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 10/12/2012

Number 220360

Expires 10/12/2014

MO 680-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)