



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-654-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DESOTO POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
04/30/14

ARREST TIME: 21:15
SUBJECT NAME:
RFI/TEST
DOB: 01/01/80 SEX: M
STATE/D.L.: MO/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: 000
TESTING OFFICER:
SNODGRASS/DANIEL/SCOTT
OFFICER I.D.: 607
PERMIT NUMBER: 220142
EXPIRATION DATE: 06/12/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:40
INTERNAL STANDARD	VERIFIED	22:40
RADIO INTERFERENCE		

Operator Signature

Pell 607 Type II

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DESOTO POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
04/30/14

ARREST TIME: 21:15
SUBJECT NAME:
SELF/TEST
DOB: 01/01/80 SEX: M
STATE/D.L.: MO/NA
ARRESTING OFFICER:
NA
OFFICER I.D.: 000
TESTING OFFICER:
SNODGRASS/DANIEL/SCOTT
OFFICER I.D.: 607
PERMIT NUMBER: 220142
EXPIRATION DATE: 06/12/14
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	22:44
INTERNAL STANDARD	VERIFIED	22:44
SUBJECT SAMPLE	.000	22:44
BLANK TEST	.000	22:45

Operator Signature

Pell 607 Type II

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DESOTO POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
04/30/14
22:32

---- DIAGNOSTIC CHECK ----

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~

Operator Signature *Roll 607 Type II*

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
DESOTO POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201208
04/30/14

TESTING OFFICER:
SNODGRASS/DANIEL/SCOTT
OFFICER I.D.: 607
PERMIT NUMBER: 220142
EXPIRATION DATE: 06/12/14
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	22:35
INTERNAL STANDARD	VERIFIED	22:35
EXTERNAL STANDARD	.101	22:36
BLANK TEST	.000	22:36
EXTERNAL STANDARD	.102	22:37
BLANK TEST	.000	22:37
EXTERNAL STANDARD	.102	22:38
BLANK TEST	.000	22:38

N = 3
SIM. = .1
AVG. = .1016

Operator Signature *Roll 607 Type II*

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DANIEL SCOTT SNODGRASS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 06/12/2012

Number 220142

Expires 06/12/2014

MO 580-6771 (7-86)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-86)