



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #8
By Carol Day at 12:08 pm, Feb 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 20129	NAME OF AGENCY PERRYVILLE PD	DATE OF INSPECTION 01/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 120 N. JACKSON ST. PERRYVILLE		TIME OF INSPECTION 10:44 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01-28-2014 10:44</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>+48 °C</u>	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO LOT # 13002 EXP. DATE 06/19/2015

SIMULATOR TEMP ($34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$) +34 °C SIMULATOR SN SD2778 EXP. DATE 09/16/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within $\pm 5\%$ of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 \rightarrow .098	TEST 2 \rightarrow .100	TEST 3 \rightarrow .100
---------------------------	---------------------------	---------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is working within DOHSS guidelines

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME WESLEY BELL
TYPE II PERMIT NUMBER/EXPIRATION DATE 230197 09/19/2015	TELEPHONE NUMBER (573) 547-4546

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS**

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

Face This Side Down - This Edge Is First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
01/28/14

ARREST TIME: 18:00
SUBJECT NAME:
DOE/JOHN/K

DOB: 04/12/77 SEX: M

STATE/D.L.: MO/123456789

ARRESTING OFFICER:
BELL/MESELEY/J

OFFICER I.D.: 119
TESTING OFFICER:
S/RN/E

OFFICER I.D.: S/RN/E

PERMIT NUMBER: 230192

EXPIRATION DATE: 03/19/15

MISCELLANEOUS DATA:
N/A
N/A

--- BREATH ANALYSIS ---

BLANK TEST .0000 18:55
INTERNAL STANDARD VERIFIED
RADIO INTERFERENCE 18:55

Operator Signature

Col. [Signature]
119

BAC DataMaster Evidence Ticket

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 201209

01/28/14

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 201209

01/28/14

10:44

TESTING OFFICER:

BELL/WESLEY/J

OFFICER I.D.: 119

PERMIT NUMBER: 230197

EXPIRATION DATE: 09/19/15

SCHEMATIC DATA:

N/A

N/A

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 4Sc

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:47
INTERNAL STANDARD	VERIFIED	10:47
INTERNAL STANDARD	.098	10:47
BLANK TEST	.000	10:48
INTERNAL STANDARD	.100	10:48
BLANK TEST	.000	10:49
INTERNAL STANDARD	.100	10:49
BLANK TEST	.000	10:50

n = 3

RM = .1

SD = .0993

PRINTER TEST:

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~

Signature

2208-02

Operator Signature

2208-02



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

WESLEY J BELL

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service, and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE: 9/19/2013

NUMBER 230197

EXPIRES 9/19/2015

Wanda
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gael Vesterly
acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-S&D-0721 (6/03)

LAB-278(10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BELL, WESLEY
Permit No 230197
Date Issued 9/19/2013 Date Expires 9/19/2016