



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 2:55 pm, May 19, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201209	NAME OF AGENCY Perryville PD	DATE OF INSPECTION 05/13/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 120 N. Jackson St., Perryville		TIME OF INSPECTION 8:04 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 05/13/2014 20:04
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER +48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repro LOT # 13002 EXP. DATE 08/19/2015

SIMULATOR TEMP (34°C ± 0.2°C) +34 °C SIMULATOR SN sd2778 EXP. DATE 09/16/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 \blacktriangleright .099	TEST 2 \blacktriangleright .100	TEST 3 \blacktriangleright .101
-----------------------------------	-----------------------------------	-----------------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1	(0-.04)	(.05-.09)	(.10-.14) 1	(.15-.19)	OVER .19
------------	---------	-----------	-------------	-----------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is working within DOHSS guidelines

INSPECTING OFFICER

SIGNATURE <i>Wesley J. Bell #119</i>	PRINT FULL NAME Wesley J. Bell
TYPE II PERMIT NUMBER EXPIRATION DATE 230197 09/19/2015	TELEPHONE NUMBER (573) 547-4546

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

**BAC DataMaster
Evidence Ticket**

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
05/13/14

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
05/13/14
20:04

TESTING OFFICER:
BELL/WESLEY/J
OFFICER I.D.: 119
PERMIT NUMBER: 230197
EXPIRATION DATE: 03/19/15
MISCELLANEOUS DATA:
N/A
N/A

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 43C
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

--- SUPERVISOR MODE ---

BLANK TEST .000 20:06
INTERNAL STANDARD VERIFIED 20:06
EXTERNAL STANDARD .099 20:07
BLANK TEST .000 20:07
EXTERNAL STANDARD .100 20:08
BLANK TEST .000 20:08
EXTERNAL STANDARD .101 20:09
BLANK TEST .000 20:09

1 = 3
SIM. = .1
WG. = .1

PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

Operator Signature Cpt. Wesley Bell #119

Operator Signature Cpt. Wesley Bell #119

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
05/13/14

ARREST TIME: 19:30
SUBJECT NAME:
DOE/JON/L
DOB: 09/21/66 SEX: M
STATE/D.L.: MO/345678904
ARRESTING OFFICER:
BELL/WESLEY/J
OFFICER I.D.: 119
TESTING OFFICER:
SAME
OFFICER I.D.: SAME
PERMIT NUMBER: 230197
EXPIRATION DATE: 09/19/15
MISCELLANEOUS DATA:
N/A
N/A

--- BREATH ANALYSIS ---

BLANK TEST .000 20:14
INTERNAL STANDARD VERIFIED 20:14
RADIO INTERFERENCE

Operator Signature *Cpt. Wesley Bell #119*



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

WESLEY J BELL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 572.020 through 572.044, RSMo and 506.111 through 506.119 RSMo.

DATE: 9/19/2013

NUMBER 230197

EXPIRES 9/19/2015

[Signature]

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-ABO-0771(6/10)

LAB-0771(6/10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BELL, WESLEY
Permit No 230197
Date Issued 9/19/2013 Date Expires 9/19/2015