STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 2:43 pm, Nov 18, 2014
REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201209	Perryville PD	11/12/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
120 N. Jackson St. Perryville	11:46 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 11/12/2014 11:46

☑ COMPUTER ☑ DETECTOR

☑ PROGRAM ☑ FILTERS

☑ HEATERS SAMPLE CHAMBER +48 °C ☑ QUARTZ STANDARD

☑ FLOW DETECTOR ☑ CALIBRATION

☑ PUMP HIGH SPEED ☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER Repco LOT # 14001 EXP. DATE 04/30/2016

☑ SIMULATOR TEMP (34°C ± 0.2°C) +34 °C SIMULATOR SN sd2778 EXP. DATE 11/06/2015

☑ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .100	TEST 2 ☛ .099	TEST 3 ☛ .100

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is working within DOHSS Guidlines

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
[signature] #119	Wesley Bell
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230197 09/19/2015	(573) 547-4546

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **14001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1216** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **May 1, 2014** The expiration date for this lot number is **April 30, 2016** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
11/12/14
11:46

--- DIAGNOSTIC CHECK ---

MPUTER:	OKAY
08RAM (04-07-2009):	OKAY
HTERS	
MPLE CHAMBER:	48C
OW DETECTOR:	OKAY
MP	
IGH SPEED:	OKAY
ETECTOR:	OKAY
ILTERS:	OKAY
ARTZ STANDARD:	OKAY
LIBRATION:	OKAY

PRINTER TEST
"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
JKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Signature Cpl. Wesley Bell #119

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
11/12/14

TESTING OFFICER:
BELL/WESLEY/J
OFFICER I.D.: 119
PERMIT NUMBER: 230137
EXPIRATION DATE: 03/19/15
MISCELLANEOUS DATA:
N/A
N/A

--- SUPERVISOR MODE ---

BLANK TEST	.000	11:49
INTERNAL STANDARD	VERIFIED	11:50
EXTERNAL STANDARD	.100	11:50
BLANK TEST	.000	11:51
EXTERNAL STANDARD	.099	11:51
BLANK TEST	.000	11:52
EXTERNAL STANDARD	.100	11:52
BLANK TEST	.000	11:53

N = 3
SIM. = .1
AVG. = .0996

Operator Signature Cpl. Wesley Bell #119

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201209
11/12/14

ARREST TIME: 11:00
SUBJECT NAME:
DOE/JON/L
DOB: 03/12/77 SEX: M
STATE/D.L.: MO/S009785432
ARRESTING OFFICER:
BELL/WESLEY/J
OFFICER I.D.: 119
TESTING OFFICER:
SAME
OFFICER I.D.: SAME
PERMIT NUMBER: 230197
EXPIRATION DATE: 09/19/15
MISCELLANEOUS DATA:
N/A
N/A

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:57
INTERNAL STANDARD	VERIFIED	11:57
RADIO INTERFERENCE		

Operator Signature Cpl. Wesley Bell #119

2208-02

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

WESLEY J BELL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/19/2013

NUMBER 230197

EXPIRES 9/19/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-9721 (6-10) LAB 4 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BELL, WESLEY
Permit No 230197
Date Issued 9/19/2013 Date Expires 9/19/2015