

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
 DATAMASTER MAINTENANCE REPORT

REPORT #

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

RECEIVED
 By Carol Day at 3:27 pm, Jun 23, 2014

DATAMASTER SN 201214	NAME OF AGENCY HAYTI Police Dept.	DATE OF INSPECTION 6-22-14
LOCATION OF INSTRUMENT (ADDRESS AND CITY) 300 EAST MAIN STREET, HAYTI		TIME OF INSPECTION 8:34 AM

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout)	6-22-14 8:34 AM
<input type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input type="checkbox"/> FILTERS	
<input type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input type="checkbox"/> QUARTZ STANDARD	
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	

<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Repro	LOT # 13002 EXP. DATE 06-19-15
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.00 °C	SIMULATOR SN 503503 EXP. DATE 07/10/14

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 * .097	TEST 2 * .098	TEST 3 * .098
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 7	(0-.04) 7	(.05-.09) 7	(.10-.14) 7	(.15-.19) 7	OVER 19 7
-------------------	------------------	--------------------	--------------------	--------------------	------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER David Mallin	PHONE # 573-359-1259
OFFICE PHONE # 240239	TELEPHONE # 05-13-16

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

REP CO MARKETING INC.

4-27-14
RECEIVED

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-9480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013.
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
MAYTTE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201214

08/29/14

08:34

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 50C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

0123456789+-= @ABCDEFGHI

!JKLMNOPQRSTUWXYZ[]^_`abc

~!@#%&'()*+,-./:;<=>?@

OPERATOR SIGNATURE



CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-528-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
HAYTI POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER: 26121A
06/22/14

TESTING OFFICER:
MACLIN, DAVID
OFFICER I.D.: 0506
PERMIT NUMBER: 240239
EXPIRATION DATE: 05/13/16
MISCELLANEOUS DATA:

--- SUPERVISOR MADE ---

BLANK TEST	000	08:40
INTERNAL STANDARD	VERIFIED	08:40
EXTERNAL STANDARD	097	08:41
BLANK TEST	000	08:41
EXTERNAL STANDARD	098	08:42
BLANK TEST	000	08:42
EXTERNAL STANDARD	098	08:43
BLANK TEST	000	08:43

SUM =
AVG = 0076

OPERATOR SIGNATURE 

CARD STK #
60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (N.P.A.S.)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
HAYTI POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201214
06/22/14

ARREST TIME: 09:30

SUBJECT NAME:

TEST

DOB: 11/11/11 SEX: M

STATE D.L. F. NO: 11111111

ARRESTING OFFICER:

MAGLIN DAVID

OFFICER I.D. #: 5061

TESTING OFFICER:

MAGLIN DAVID

OFFICER I.D. #: 506

PERMIT NUMBER: 24023

EXPIRATION DATE: 05/13/16

MISCELLANEOUS DATA:

BREATH ANALYSIS

BLAN TEST: 03/16

INTERNAL STANDARD: 018/16

RATIO REFERENCE:

OPERATOR SIGNATURE



CARD STK #
80038

REORDER ALL SUPPLIES FROM N.P.A.S.
2280 NORTH MAIN, MANSFIELD, OH 44905-416, 520-6727 (N.P.A.S.)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

DAVID MACLIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/13/2014


NUMBER 240239

EXPIRES 5/13/2016

MO 580-0771 (6-10)



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY



acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **MACLIN, DAVID**
Permit No **240239**
Date Issued **5/13/2014** Date Expires **5/13/2016**