



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
 By Carol Day at 12:23 pm, Nov 24, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN #201217	NAME OF AGENCY Crystal City Police Department	DATE OF INSPECTION 11/23/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 130 Mississippi Avenue, Crystal City		TIME OF INSPECTION 2:30 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11/23/2014 14:56 hours
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 14200 EXP. DATE 08/05/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN DR3772 EXP. DATE 04/30/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099	TEST 2 .099	TEST 3 .098
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Sgt. Jeffery S. Wynn, DSN-204
TYPE II PERMIT NUMBER/EXPIRATION DATE 240141 04/03/2016	TELEPHONE NUMBER (636) 937-4601

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

890 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-644-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FNI22211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CRYSTAL CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201217
11/23/14
14:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 48c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~

Operator Signature

[Signature] #204 Type II

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CRYSTAL CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201217
11/23/14

TESTING OFFICER:
WYNN/JEFFERY/S
OFFICER I.D.: 204
PERMIT NUMBER: 240141
EXPIRATION DATE: 04/03/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 14:59
INTERNAL STANDARD VERIFIED 14:59
EXTERNAL STANDARD .000 15:00
BLANK TEST .000 15:01
EXTERNAL STANDARD .000 15:01
BLANK TEST .000 15:02
EXTERNAL STANDARD .000 15:02
BLANK TEST .000 15:03

H = 9
SIG. = .1
AVG. = .0996

Operator Signature

[Signature] #204 Type II

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
CRYSTAL CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201217
11/23/14

ARREST TIME: 14:30
SUBJECT NAME:
TEST/RF1
DOB: 01/02/84 SEX: M
STATE/D.L.: MO/A123456789
ARRESTING OFFICER:
WYNN/JEFFERY/S
OFFICER I.D.: 204
TESTING OFFICER:
WYNN/JEFFERY/S
OFFICER I.D.: 204
PERMIT NUMBER: 240141
EXPIRATION DATE: 04/03/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.200	15:05
INTERNAL STANDARD	VERIFIED	15:06
RADIO INTERFERENCE		

Operator Signature

Sgt. Wynn #204 Type II



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JEFFERY S WYNN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014


NUMBER 240141

EXPIRES 4/3/2016

MO 680-0771 (6-10)



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY



DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator WYNN, JEFFERY
Permit No 240141
Date Issued 4/3/2014 Date Expires 4/3/2016