



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 10:48 am, May 28, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201219	NAME OF AGENCY Kearney Police Department	DATE OF INSPECTION 05/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 725 West 92 Highway Kearney, Missouri, 64060		TIME OF INSPECTION 10:00 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>05/28/2014 10:00</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ <u>49</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc LOT # 13100 EXP. DATE 04/23/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2751 EXP. DATE 12/17/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .104	TEST 2 .104	TEST 3 .104
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	4	(0-.04)	0	(.05-.09)	1	(.10-.14)	0	(.15-.19)	1	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument operated within established limits.

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Derick E. Hager
TYPE II PERMIT NUMBER/EXPIRATION DATE 240256 / 05/20/2016	TELEPHONE NUMBER (816) 903-4768

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201219
05/28/14

ARREST TIME: 09:45
SUBJECT NAME:
RFICHECK

DOB: 10/18/81 SEX: M
STATE/D.L.: MO/1234567890

ARRESTING OFFICER:
HAGER/DERICK/E

OFFICER I.D.: 0981
TESTING OFFICER:
HAGER/DERICK/E

OFFICER I.D.: 0981
PERMIT NUMBER: 240256

EXPIRATION DATE: 05/20/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 10:10
INTERNAL STANDARD VERIFIED 10:10
RADIO INTERFERENCE

OPERATOR SIGNATURE
Card Stock No.
60021

Derick Hager #981

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201219
05/28/14
10:00

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS
SAMPLE CHAMBER: 49C

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST
!"###\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~

OPERATOR SIGNATURE
Card Stock No.
60021

Derick Hager #981

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201219
05/28/14

ARREST TIME: 10:00
SUBJECT NAME: ICE/JOHN
DOB: 01/10/91 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER: HAGER/DERICK/E
OFFICER I.D.: 0981
TESTING OFFICER: HAGER/DERICK/E
OFFICER I.D.: 0981
PERMIT NUMBER: 240256
EXPIRATION DATE: 05/20/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:18
INTERNAL STANDARD	VERIFIED	10:18
SUBJECT SAMPLE	.000	10:19
BLANK TEST	.000	10:19

OPERATOR SIGNATURE
Card Stock No. 60021

Devil Log #197

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 201219
05/28/14

TESTING OFFICER: HAGER/DERICK/E
OFFICER I.D.: 0981
PERMIT NUMBER: 240256
EXPIRATION DATE: 05/20/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:13
INTERNAL STANDARD	VERIFIED	10:13
EXTERNAL STANDARD	.104	10:13
BLANK TEST	.000	10:14
EXTERNAL STANDARD	.104	10:14
BLANK TEST	.000	10:15
EXTERNAL STANDARD	.104	10:15
BLANK TEST	.000	10:16

N = 3
SIM. = .1
AVG. = .104

OPERATOR SIGNATURE
Card Stock No. 60021

Devil Log #198



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DERICK E HAGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/20/2014

NUMBER 240256

EXPIRES 5/20/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HAGER, DERICK
Permit No 240256
Date Issued 5/20/2014 Date Expires 5/20/2016