RECEIVED
By Carol Day at 9:07 am, Feb 06, 2014

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER DMT S/N	NAME OF AGENCY	DATE OF INSPECTION
201220	Scott city Police Department	2-5-2014
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
618 Main street Scott city		07:27

CHECKLIST: Place a mark on the line by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

- [X] DIAGNOSTIC RECORD
  - DATE AND TIME 2-5-14 07:27
  - [X] PROGRAM
  - [X] SAMPLE CHAMBER
  - [X] BREATH TUBE
  - [X] PUMP
  - [X] DETECTOR
  - [X] FILTER 1
  - [X] FILTER 2
  - [X] FILTER 3
  - [X] INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

- [X] SIMULATOR SOLUTION
- [ ] COMPRESSED ETHANOL-GAS MIXTURE
- [X] STANDARD SUPPLIER Repco Marketing LOT # 12002 EXP. DATE 8-29-2014
- [X] SIMULATOR TEMP (34°C ± 0.2°C) 34° C SIMULATOR S/N SP2220 SIMULATOR EXP DATE
- [X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.
  - [X] 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
  - [ ] 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
  - [ ] 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: .097	TEST 2: .097	TEST 3: .097

- [X] PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TEST IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS 2	0-.04 0	.05-.09 0	.10-.14 2	.15-.19 0	OVER .19 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Dan McMichael	Dan McMichael
TYPE II PERMIT NUMBER: 210104	EXPIRATION DATE: 01-03-2015
TELEPHONE NUMBER: 573-264-2121	

RETURN COMPLETED REPORT TO THE
Breath Alcohol Program, Missouri Department of Health and Senior Services, Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901

MO 580-2838 (7-10) AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER services provided on a nondiscriminatory basis LAB-160

02/05/2014 08:10 5762642125 SCOTT CITY POLICE PAGE 01/05

REPCO MARKETING INC.

3101-180 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-878-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 12002
EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012 The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
02/05/14

ARREST TIME: 06:50
SUBJECT NAME:
MCMICHAEL/DAN/LEE
DOB: 10/10/50 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	07:36
INTERNAL STANDARD	VERIFIED	07:36
RADIO INTERFERENCE		

OPERATOR SIGNATURE

CARD STK # 60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
02/05/14

ARREST TIME: 06:50
SUBJECT NAME:
MCMICHAEL/DAN/LEE
DOB: 10/10/50 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	07:38
INTERNAL STANDARD	VERIFIED	07:38
SUBJECT SAMPLE	.000	07:39
BLANK TEST	.000	07:39

OPERATOR SIGNATURE

CARD STK # 60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
02/05/14
07:27

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

OPERATOR SIGNATURE

CARD STK # 60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
02/05/14

TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 103
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	07:30
INTERNAL STANDARD	VERIFIED	07:30
EXTERNAL STANDARD	.097	07:31
BLANK TEST	.000	07:31
EXTERNAL STANDARD	.097	07:32
BLANK TEST	.000	07:32
EXTERNAL STANDARD	.097	07:33
BLANK TEST	.000	07:33

N = 3
SIM. = .1
AVG. = .097

OPERATOR SIGNATURE

CARD STK # 60036

REORDER ALL SUPPLIES FROM N.P.A.S.
2260 NORTH MAIN, MANSFIELD, OH 44903 419-526-6727 (NPAS)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DAN MCMICHAEL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/03/2013

NUMBER 210104

EXPIRES 01/03/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (9-10)

LAB-4 (Rev-10)