RECEIVED
By Carol Day at 12:47 pm, Jul 03, 2014

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER DMT SN	NAME OF AGENCY	DATE OF INSPECTION
201220	Scott city Police Department	6-28-14
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
618 Main street	scott city	1514

CHECKLIST: Place a mark on the line by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC RECORD
  - DATE AND TIME 6-28-14 - 1514
  - [x] PROGRAM
  - [x] SAMPLE CHAMBER 49° C
  - [x] BREATH TUBE
  - [x] PUMP
  - [x] DETECTOR
  - [x] FILTER 1
  - [x] FILTER 2
  - [x] FILTER 3
  - [x] INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

- [x] SIMULATOR SOLUTION
- [ ] COMPRESSED ETHANOL-GAS MIXTURE
- [x] STANDARD SUPPLIER Repco Marketing LOT # 13002 EXP. DATE 6-19-15
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34°C SIMULATOR SN 52230 SIMULATOR EXP DATE
- [x] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

- [x] 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1:	TEST 2:	TEST 3:
.096	.096	.095

- [x] PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TEST IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS	0-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	1	0	1	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE	PERMIT FULL NAME
Dan McMichael	Dan McMichael
TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
210104	01-03-2015	573-264-2121

RETURN COMPLETED REPORT TO THE
Breath Alcohol Program, Missouri Department of Health and Senior Services, Southeast District Office, 2875 James Blvd, Poplar Bluff, MO 63901

MO 580-2898 (7-10) AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER LAB-168

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, NC. 27604
919-876-5430

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1217** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **June 20, 2013**
The expiration date for this lot number is **June 19, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
06/28/14
15:14

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature Dan McMichael

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
06/28/14

TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	15:28
INTERNAL STANDARD	VERIFIED	15:28
EXTERNAL STANDARD	.096	15:28
BLANK TEST	.000	15:29
EXTERNAL STANDARD	.096	15:29
BLANK TEST	.000	15:30
EXTERNAL STANDARD	.095	15:30
BLANK TEST	.000	15:31

N = 3
STM. = .1
AVG. = .0956

Operator Signature Dan McMichael

2208-0

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201220
06/28/14

ARREST TIME: 14:30
SUBJECT NAME:
BUNDY/AL
DOB: 10/10/50 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	15:38
INTERNAL STANDARD	VERIFIED	15:38
SUBJECT SAMPLE	.000	15:39
BLANK TEST	.000	15:39

Operator Signature

2208-02

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SCOTT CITY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201320
06/28/14

ARREST TIME: 14:30
SUBJECT NAME:
BUNDY/AL
DOB: 10/10/50 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
TESTING OFFICER:
MCMICHAEL/DAN/LEE
OFFICER I.D.: 105
PERMIT NUMBER: 210104
EXPIRATION DATE: 01/03/15
MISCELLANEOUS DATA:
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	15:36
INTERNAL STANDARD	VERIFIED	15:36
RADIO INTERFERENCE		

Operator Signature

2206-02

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DAN MCMICHAEL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 01/03/2013

NUMBER 210104

EXPIRES 01/03/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-7 (06-10)