MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 3:09 pm, Jan 24, 2014

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201223	Branson West police Department	01/20/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
110 Silverlady Ln Branson West Mo, 65737	6:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 01-20-2014 1800
- ☑ COMPUTER — ☑ DETECTOR
- ☑ PROGRAM — ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 49 °C — ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR — ☑ CALIBRATION
- ☑ PUMP HIGH SPEED — ☑ PRINTER
- ☑ INDICATOR LIGHTS
- ☑ SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13100 EXP. DATE 04/29/2015
- ☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3139 EXP. DATE 08/01/2014
- ☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛	TEST 2 ☛	TEST 3 ☛
.099	.099	.100

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	2	0	2	3	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸ 	Vincent Collins
TYPE II PERMIT NUMBER/EXPIRATION DATE: 230091 05/28/2015	TELEPHONE NUMBER: (417) 272-3400

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscrimatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111- 4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13100** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 29, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 23, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN122211-02** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

VINCENT COLLINS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 05/28/2013

NUMBER 230091

EXPIRES 05/28/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BRANSON WEST POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201223
01/20/14
18:00

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}+~
```

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BRANSON WEST POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201223
01/20/14

TESTING OFFICER:
COLLINS/VINCENT/T
OFFICER I.D.: 383
PERMIT NUMBER: 230091
EXPIRATION DATE: 05/28/15
MISCELLANEOUS DATA:
N
N

--- SUPERVISOR MODE ---

BLANK TEST	.000	18:03
INTERNAL STANDARD	VERIFIED	18:03
EXTERNAL STANDARD	.099	18:04
BLANK TEST	.000	18:05
EXTERNAL STANDARD	.099	18:05
BLANK TEST	.000	18:06
EXTERNAL STANDARD	.100	18:06
BLANK TEST	.000	18:07

N = 3
SIM. = .1
AVG. = .0993
S.D. = .0005
S.E. = .6666

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BRANSON WEST POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201223
01/20/14

ARREST TIME: 17:50
SUBJECT NAME:
RFI/TEST
DOB: 01/01/80 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
COLLINS/VINCENT/T
OFFICER I.D.: 303
TESTING OFFICER:
COLLINS/VINCENT/T
OFFICER I.D.: 303
PERMIT NUMBER: 230091
EXPIRATION DATE: 05/28/15
MISCELLANEOUS DATA:
N
N

--- BREATH ANALYSIS ---

BLANK TEST	.000	18:10
INTERNAL STANDARD	VERIFIED	18:10
RADIO INTERFERENCE		

Operator Signature

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BRANSON WEST POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201223
01/20/14

ARREST TIME: 17:50
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/80 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
COLLINS/VINCENT/T
OFFICER I.D.: 303
TESTING OFFICER:
COLLINS/VINCENT/T
OFFICER I.D.: 303
PERMIT NUMBER: 230091
EXPIRATION DATE: 05/28/15
MISCELLANEOUS DATA:
N
N

--- BREATH ANALYSIS ---

BLANK TEST	.000	18:20
INTERNAL STANDARD	VERIFIED	18:20
SUBJECT SAMPLE	.000	18:20
BLANK TEST	.000	18:21

Operator Signature