



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

ORIGINAL

RECEIVED
 By Carol Day at 3:57 pm, Aug 12, 2014
 REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201230	NAME OF AGENCY Cape Girardeau Police Department	DATE OF INSPECTION 08/02/2014
-------------------------	----------------------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 40 South Sprigg Street, Cape Girardeau, MO 63703 (Booking)	TIME OF INSPECTION 8:14 am
--------------------------------------------------------------------------------------------------------	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 08-02-2014 08:14
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Repro LOT # 13802 EXP. DATE 08/12/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2221 EXP. DATE 07/14/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .079	TEST 2 .080	TEST 3 .080
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	48	(.05-.09)	4	(.10-.14)	0	(.15-.19)	5	OVER .19	2
----------	---	---------	----	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument was found to be within DOH Specs

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Jeffrey D Lucas
TYPE II PERMIT NUMBER/EXPIRATION DATE 240223 05/10/2016	TELEPHONE NUMBER (573) 335-6621

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

RepCo Marketing Inc.

NO-18 STANDARD FORM
MAY 1962 EDITION
GPO: 1962-0-242-000

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13802

EXPIRATION DATE: August 12, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13802 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .0968 gms/dl $\pm .003$ gms/dl w/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .080 $\pm 3\%$ gms/210L Breath when heated to 34 Degrees Celsius ± 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 13, 2013.
The expiration date for this lot number is August 12, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JEFFREY D LUCAS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/10/2014

NUMBER 240223

EXPIRES 5/10/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MSD 680-0771 (P-10)

LAS-4 (P15-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator LUCAS, JEFFREY
 Permit No 240223
 Date Issued 5/10/2014 Date Expires 5/10/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS
1000 WEST WASHINGTON
LANSING, MICHIGAN 48207

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS

CURRENTLY: YES NO

PREVIOUSLY USED FOR OTHER: YES NO

NOTED: YES NO

SAMPLE CONTAINER: YES NO

FLASK MATERIAL: YES NO

PLUG: YES NO

HIGH SPEED: YES NO

INJECTOR: YES NO

FIELDKIT: YES NO

QUARTZ STANDARD: YES NO

CHLORAMPHENICOL: YES NO

QUANTITY: 12.00

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS
1000 WEST WASHINGTON
LANSING, MICHIGAN 48207

Operator Signature

J. Lucas #226

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MICHIGAN

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS
1000 WEST WASHINGTON
LANSING, MICHIGAN 48207

TESTING OF SAMPLES:

1. INJECTOR: YES NO

2. PLUG: YES NO

3. HIGH SPEED: YES NO

4. FIELDKIT: YES NO

5. QUARTZ STANDARD: YES NO

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS

TEST	DATE	BY
INJECTOR	10/12/00	DL
PLUG	10/12/00	DL
HIGH SPEED	10/12/00	DL
FIELDKIT	10/12/00	DL
QUARTZ STANDARD	10/12/00	DL
CHLORAMPHENICOL	10/12/00	DL
BLANK TEST	10/12/00	DL

LABORATORY FOR THE DEPARTMENT OF CORRECTIONS
1000 WEST WASHINGTON
LANSING, MICHIGAN 48207

Operator Signature

J. Lucas #226

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI
LABORATORIES (POLICE DEPARTMENT)

LABORATORY REFERENCE NUMBER: 10040278

REMIK, LORENE

SUBJECT: MURDER

OFFENSE: MURDER

DOB: 01/10/60

STATE: D.C. NO. 00000

ARRESTING OFFICER:

OFFENSE:

OFFICER: J. D. J. BUN

TESTING OFFICER:

LUCAS JEFFREY

OFFICER: J. D. J. BUN

PERMIT NUMBER: 00000

EXPIRATION DATE: 05/10/10

MISCELLANEOUS DATA:

LABORATORY NUMBER:

BLANK: YES

INTERNAL STANDARD

REF ID: A111111111

DATE: 05/10/10
TIME: 09:00

Operator Signature

J Lucas #059