MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 2:55 pm, Dec 18, 2014

REPORT # 6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201232	Salem Police Department	12/08/2014

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
500 North Jackson Street, Salem 65560	8:47 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 12/08/2014 20:47
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 49 °C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 14200 EXP. DATE 08/05/2016
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2774 EXP. DATE 06/04/2015
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.098	.099	.099

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	1	2	0	1

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Sgt. - Robert T Simpson #105	Robert T. Simpson
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240340 09/22/2016	(573) 729-4242

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14200** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **August 6, 2014,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1213%** (w/vol) ethyl alcohol. The expiration date for this lot number is **August 5, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SALEM POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201232
12/08/14

ARREST TIME: 20:30
SUBJECT NAME:
[illegible]
DOB: 11/11/11 SEX: M
[illegible]: NA/NA
ARRESTING OFFICER:
[illegible]
OFFICER I.D.: NA
TESTING OFFICER:
[illegible]/T
OFFICER I.D.: 105
[illegible] NUMBER: 240340
EXPIRATION DATE: 09/22/16
MISCELLANEOUS DATA:
[illegible]

--- BREATH ANALYSIS ---

[illegible] TEST	.000	20:55
[illegible] STANDARD	VERIFIED	20:56
[illegible] INTERFERENCE		

Operator Signature Sgt. Robert T Simpson #105

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SALEM POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201232
12/08/14
20:47

---- DIAGNOSTIC CHECK ---

[illegible]	OKAY
[illegible] (04-07-2009):	OKAY
[illegible]	
[illegible]	49c
[illegible]	OKAY
[illegible]	
[illegible]	OKAY
[illegible]	OKAY
[illegible]	OKAY
[illegible] STANDARD:	OKAY
[illegible]	OKAY

PRINTER TEST
[illegible] 0123456789:;<=>?@ABCDEFG
[illegible] abcdefghijklmno
[illegible]

Operator Signature Sgt. Robert J Simpson #105

Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SALEM POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201232
12/08/14

TESTING OFFICER:
[illegible]/ROBERT/T
OFFICER I.D.: 105
TEST NUMBER: 240840
CALIBRATION DATE: 09/28/16
[illegible] DATA:

---- SUPERVISOR MODE ---

[illegible] TEST	.000	20:50
[illegible] STANDARD	VERIFIED	20:50
[illegible] STANDARD	.098	20:51
[illegible] TEST	.000	20:51
[illegible] STANDARD	.099	20:52
[illegible] TEST	.000	20:52
[illegible] STANDARD	.099	20:53
[illegible] TEST	.000	20:53

[illegible] = [illegible]
[illegible] = [illegible]
[illegible] = .0996

Operator Signature Sgt. Robert J Simpson #105

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

ROBERT T SIMPSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/22/2014

NUMBER 240340

EXPIRES 9/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling ,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)