



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

RECEIVED REPORT #  
By Carol Day at 9:20 am, Feb 06, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201233	NAME OF AGENCY BRENTWOOD POLICE	DATE OF INSPECTION 02/02/2014			
LOCATION OF INSTRUMENT (STREET AND CITY) 272 HANLEY INDUSTRIAL COURT BRENTWOOD MO 63144		TIME OF INSPECTION 11:30 pm			
<b>CHECKLIST:</b> Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.					
<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)		DATE AND TIME (from printout) 02/02/14 23:30			
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR				
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS				
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD				
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION				
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER				
<input checked="" type="checkbox"/> INDICATOR LIGHTS					
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER REPCO MARKETING INC. LOT # 13002 EXP. DATE 06/19/2015					
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2762 EXP. DATE 02/14/2014					
<input checked="" type="checkbox"/> CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)					
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)					
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE					
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE					
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE					
TEST 1 → .097	TEST 2 → .098	TEST 3 → .098			
<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)					
INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)					
REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 0
LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).					
<b>INSPECTING OFFICER</b>					
SIGNATURE ▶			PRINT FULL NAME MATT MORGAN		
TYPE II PERMIT NUMBER/EXPIRATION DATE 230228 10/17/2015			TELEPHONE NUMBER (314) 644-7100		
RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office 2875 James Blvd. Poplar Bluff, MO 63901					

*REPCO MARKETING INC.*

3101-188 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-878-5480

### CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSISSIPPI  
BRENTWOOD POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER: 1000000000  
02/25/14  
08:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OK

PROGRAM (04-27-2008): OK

HEATERS: OK

SAMPLE VOLUMES: OK

FLOW DETECTOR: OK

PUMP: OK

HIGH SPEED: OK

DETECTORS: OK

FILTERS: OK

QUARTZ STANDARD: OK

CALIBRATION: PASS

PRINTER TEST: OK

#6667 ( ) 04-27-2008  
HIGH SENSITIVITY OVER 20% IN 10 SECONDS IN 100  
% SENSITIVITY

Operator Signature: *[Signature]*

230218

10-17-15

Face This Side Down - This Edge In First

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSISSIPPI  
BRENTWOOD POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 251138  
02/22/14

ARREST TIME: 22:00

SUBJECT NAME:

RFI/TEST

DOB: 01/01/81 SEX: M

STATE/CLASS: MD/

ARRESTING OFFICER:

FORGAN

OFFICER I.D.#: 150

TESTING OFFICER:

MURPHY

OFFICER I.D.#: 150

PERMIT NUMBER: 223626

EXPIRATION DATE: 12/17/13

MISCELLANEOUS DATA:

RFI/TEST

--- BREATH ANALYSIS ---

BLANK TEST

INTERNAL STANDARD

RADIO INTERFERENCE

2008 23:30

VERIFIED 23:30

Operator Signature

*PO Morgan*

230272

01-21-14

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
BRENTWOOD POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201233  
02/22/14

TESTING OFFICER:

MORGAN/M

OFFICER I.D.: 156

PERMIT NUMBER: 230228

EXPIRATION DATE: 12/17/15

MISCELLANEOUS DATA:

JAN MAINT

--- SUPERVISOR MODE ---

BLANK TEST	.000	20:50
INTERNAL STANDARD	VERIFIED	20:50
EXTERNAL STANDARD	.097	20:50
BLANK TEST	.000	20:54
EXTERNAL STANDARD	.096	20:54
BLANK TEST	.000	20:55
EXTERNAL STANDARD	.096	20:55
BLANK TEST	.000	20:56

N = 3

SIM. = .1

AVG. = .0975

OPERATOR SIGNATURE *[Signature]*

CARD STK #  
60036

REORDER ALL SUPPLIES FROM N.P.A.S. 230228  
P.O. BOX 1435, MANSFIELD, OH 44901

10-17-15



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**MATTHEW L MORGAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2013

NUMBER 230228

EXPIRES 10/17/2015

\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

\_\_\_\_\_  
acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 590-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator MORGAN, MATTHEW  
Permit No 230228  
Date issued 10/17/2013 Date Expires 10/17/2015