



19 CSR 25-30—DEPARTMENT OF HEALTH AND SENIOR SERVICES

Division 25—State Public Health Laboratory

**RECEIVED**  
By Carol Day at 11:08 am, Jan 22, 2014



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>201236</u>	NAME OF AGENCY <u>Kearney Police Department</u>	DATE OF INSPECTION <u>01-22-2014</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>725 WEST 92 Highway, Kearney, Mo, 64060</u>		TIME OF INSPECTION <u>1016</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01-22-14 @ 1016</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>48</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>GUTH LABORATORIES, Inc</u> LOT # <u>13100</u> EXP. DATE <u>04-23-2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C SIMULATOR SN <u>S02751</u> EXP. DATE <u>12-17-2014</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 → <u>.097</u>	TEST 2 → <u>.099</u>	TEST 3 → <u>.099</u>
----------------------	----------------------	----------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(.0-04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	2	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE <u>Derick E. Hager #981</u>	PRINT FULL NAME <u>DERICK E. HAGER</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>230186 / 09-04-2015</u>	TELEPHONE NUMBER <u>(816) 903-4768</u>

RETURN COMPLETED REPORT TO THE:  
Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13100 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 29, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is April 23, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
KEARNEY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201236  
01/22/14

ARREST TIME: 10:00  
SUBJECT NAME:  
RFICHECK  
DOB: 10/18/81 SEX: M  
STATE/D.L.: MO/1234567890  
ARRESTING OFFICER:  
HAGER/DERICK/E  
OFFICER I.D.: 0981  
TESTING OFFICER:  
HAGER/DERICK/E  
OFFICER I.D.: 0981  
PERMIT NUMBER: 230186  
EXPIRATION DATE: 09/04/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST .000 10:20  
INTERNAL STANDARD VERIFIED 10:20  
RADIO INTERFERENCE.

OPERATOR SIGNATURE Devil Hager #981

Card Stock No.  
60021  
REORDER ALL SUPPLIES FROM N.P.A.S.  
P.O. BOX 1465, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

**BAC DataMaster**  
Evidence Ticket

STATE OF MISSOURI  
KEARNEY POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201236  
01/22/14  
10:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM (04-07-2009): OKAY  
HEATERS  
SAMPLE CHAMBER: 48C  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!##\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrs tuvwxyz{|}~"

OPERATOR SIGNATURE Devil Hager #981

Card Stock No.  
60021  
REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

# BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
KEARNEY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201236  
01/22/14

ARREST TIME: 10:00  
SUBJECT NAME:  
HAGER/DERICK/E  
DOB: 10/18/81 SEX: M  
STATE/D.L.: MO/1234567890  
ARRESTING OFFICER:  
HAGER/DERICK/E  
OFFICER I.D.: 0981  
TESTING OFFICER:  
HAGER/DERICK/E  
OFFICER I.D.: 0981  
PERMIT NUMBER: 230186  
EXPIRATION DATE: 09/04/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	10:31
INTERNAL STANDARD	VERIFIED	10:31
SUBJECT SAMPLE	.000	10:32
BLANK TEST	.000	10:32

OPERATOR SIGNATURE *David Hager #981*

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

# BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
KEARNEY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201236  
01/22/14

TESTING OFFICER:  
HAGER/DERICK/E  
OFFICER I.D.: 0981  
PERMIT NUMBER: 230186  
EXPIRATION DATE: 09/04/15  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	10:23
INTERNAL STANDARD	VERIFIED	10:23
EXTERNAL STANDARD	.097	10:23
BLANK TEST	.000	10:24
EXTERNAL STANDARD	.099	10:25
BLANK TEST	.000	10:25
EXTERNAL STANDARD	.099	10:26
BLANK TEST	.000	10:27

N = 3  
SIM. = .1  
AVG. = .0983

OPERATOR SIGNATURE *David Hager #981*

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**DERICK E HAGER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/4/2013

NUMBER 230186

EXPIRES 9/4/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator HAGER, DERICK  
 Permit No 230186  
 Date Issued 9/4/2013 Date Expires 9/4/2015