



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

received 1/14/14-cd

REVIEWED REPORT 6
 By Carol Day at 1:42 pm, Feb 04, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201238	NAME OF AGENCY Webster Groves Police Department	DATE OF INSPECTION 01-06-2014
LOCATION OF INSTRUMENT (STREET AND CITY) 4 South Elm, Webster Groves		TIME OF INSPECTION 0844

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01-06-2014 0844
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER Guth Laboratories LOT # 13210 EXP. DATE 07-29-2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3508 EXP. DATE 02-06-2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .100	TEST 2 .101	TEST 3 .100
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	2	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument tested within normal parameters

INSPECTING OFFICER	
SIGNATURE Sgt. Jason C Flanery 2/6	PRINT FULL NAME Jason C Flanery
TYPE II PERMIT NUMBER/EXPIRATION DATE 230117 06-12-2015	TELEPHONE NUMBER 314-645-3000

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

660 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 13210 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 31, 2013, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1216% (w/vol) ethyl alcohol. The expiration date for this lot number is July 29, 2015 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

JASON C FLANERY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 06/12/2013

NUMBER 230117

EXPIRES 06/12/2015

W. W. S.

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David Voster
Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
WESTER GROVES POLICE DEPARTMENT

AND BREWSTER SERIAL NUMBER 201238
DATE: 01/26/14

LOW DETECTOR: OKAY
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
BURNZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
1#%&*()#* - \ / 0123456789: ; < (=) * ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

OPERATOR SIGNATURE *Sgt. Sean Flannery* 216

Card Stock No. 60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1495, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
WESTER GROVES POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201238
DATE: 01/26/14

TESTING OFFICER:
FLANNERY/JPSON/C
OFFICER I.D.: 216
PERMIT NUMBER: 230117
EXPIRATION DATE: 06/12/15
MISCELLANEOUS DATA:
TEST/JANUARY

BLANK TEST
INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST
EXTERNAL STANDARD
BLANK TEST
INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST

VERIFIED
08:48
08:48
08:49
08:49
08:50
08:50
08:50
08:51

N = 3
SIM. = .1
RMS. = .1000

OPERATOR SIGNATURE *Sgt. Sean Flannery* 216

Card Stock No. 60021
REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1495, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
WEBSTER GROVES POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201238
01/06/14

ARREST TIME: 08:30
SUBJECT NAME:
TEST/JANUARY/RI
DOB: 01/06/14 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
FLANERY/JRSDM/C
OFFICER I.D.: 216
TESTING OFFICER:
FLANERY/JRSDM/C
OFFICER I.D.: 216
PERMIT NUMBER: 230117
EXPIRATION DATE: 06/12/15
MISCELLANEOUS DATA:
TEST/RADIO INTER/JANUARY

--- BREATH ANALYSIS ---

--- RADIO INTERFERENCE ---

OPERATOR SIGNATURE

St. John Flanery 216

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1495, MANSFIELD, OH 44901