



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 12:26 pm, Oct 28, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201239	NAME OF AGENCY Lebanon Police Department	DATE OF INSPECTION 10/17/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 401 S Jefferson, Lebanon, Mo. 65536		TIME OF INSPECTION 2:24 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout)	10/17/2014 02:24 hrs
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER	49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR		<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED		<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER	RepCo Marketing	LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C)	34.0 °C	SIMULATOR SN SD2755 EXP. DATE 08/12/2015
<input checked="" type="checkbox"/> CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)		
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)		
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE		
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE		
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE		

TEST 1 → .096	TEST 2 → .097	TEST 3 → .097
---------------	---------------	---------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	OVER .19	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is working within DOH specifications.

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Matthew S. Wright
TYPE II PERMIT NUMBER/EXPIRATION DATE 230033 02/19/2015	TELEPHONE NUMBER (417) 532-3131

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013.
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.




Cecil B. Garner, President
RepCo Marketing, Inc.



Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19 CSR 25-30.051 (4).

Checked: 08/12/2014 Expires: 08/12/2015
Digital Therm. SN:094948
MSC Tech:DRL Temp:34.00
Agency: Lebanon Police Dept
SD 2755



Technician Printed Name: DAN LUCAS

Technician Signature: *DRL*

Date: 8/12/2014

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

MATTHEW S WRIGHT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 02/19/2013

M. S. Wright

NUMBER 230033

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

EXPIRES 02/19/2015

Shad Vukobratovic
 Acting Director

MO 580-C771 (8-10)

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RS-10)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
LABORATORY POLICE DEPARTMENT

LAB. IDENTIFICATION SERIAL NUMBER 001209
10/17/14

TESTING OF TUBES

INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST
EXTENDED STANDARD
BLANK TEST
EXTENDED STANDARD
BLANK TEST

--- SUBSEQUENT PAGES ---

INTERNAL STANDARD	0.000	0.0000
EXTERNAL STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000
EXTENDED STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000
EXTENDED STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000

N = 3
SIM. = 0.0
AVG. = 0.0000

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
LABORATORY POLICE DEPARTMENT

LAB. IDENTIFICATION SERIAL NUMBER 001209
10/17/14

--- SUBSEQUENT PAGES ---

INTERNAL STANDARD	0.000	0.0000
EXTERNAL STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000
EXTENDED STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000
EXTENDED STANDARD	0.000	0.0000
BLANK TEST	0.000	0.0000

INTERNAL STANDARD
EXTERNAL STANDARD
BLANK TEST
EXTENDED STANDARD
BLANK TEST
EXTENDED STANDARD
BLANK TEST

Operator Signature

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
LABORATORY POLICE DEPARTMENT

LABORATORY POLICE DEPARTMENT
3001 W. 11th St.
COLUMBIA, MO 65211

INVESTIGATOR: [REDACTED]

SUBJECT NAME:

[REDACTED]

DATE RECEIVED: [REDACTED]

LABORATORY USE ONLY:

ANALYST: [REDACTED]

RE:

OFFENSE CODE: [REDACTED]

TEST CODE: [REDACTED]

LABORATORY USE ONLY:

OFFENSE CODE: [REDACTED]

PERMIT NUMBER: [REDACTED]

EXPIRATION DATE: [REDACTED]

LABORATORY USE ONLY:

DRUG ANALYSIS -

BLANK TEST	_____	DATE	_____
INTERNAL STANDARD	_____	VERIFIED	_____
RADIO INTERFERENCE	_____		

Operator Signature

