



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 8:56 am, Dec 08, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201239	NAME OF AGENCY Lebanon Police Department	DATE OF INSPECTION 11/28/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 401 S Jefferson, Lebanon, Mo. 65536		TIME OF INSPECTION 9:26 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 11/28/2014 @ 09:26 hrs
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing	LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C	SIMULATOR SN SD2755 EXP. DATE 08/12/2015
<input checked="" type="checkbox"/> CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE	

TEST 1 <input checked="" type="checkbox"/> .097	TEST 2 <input checked="" type="checkbox"/> .098	TEST 3 <input checked="" type="checkbox"/> .098
---	---	---

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 Instrument is working within DOH specifications.

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Matthew S. Wright
TYPE II PERMIT NUMBER/EXPIRATION DATE 230033 02/19/2015	TELEPHONE NUMBER (417) 532-3131

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 13002

EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013
The expiration date for this lot number is June 19, 2015 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

MATTHEW S WRIGHT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATA MASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 02/19/2013

NUMBER 230033

EXPIRES 02/19/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

MS Wright
Paul Vorkel

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI
LABORATORY POLICE DEPARTMENT

LAB. DATA MASTER SERIAL NUMBER 091230
11/20/10

TESTING OFFICER:
MURPHY/RYAN/THOMAS
OFFICER ID#: 113
PERSON NUMBER: 000000
EXPIRATION DATE: 06/30/11
PERCENTAGE: 0.00%

SUPPLY NUMBER: 0000

BLANK TEST	0.00%	0.00%
INTERNAL STANDARD	0.00%	0.00%
EXTERNAL STANDARD	0.00%	0.00%
BLANK TEST	0.00%	0.00%
EXTERNAL STANDARD	0.00%	0.00%
BLANK TEST	0.00%	0.00%
EXTERNAL STANDARD	0.00%	0.00%
BLANK TEST	0.00%	0.00%

DATE: 11/20/10
SITE: 11
INV#: 100000

Operator Signature



2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI
LABORATORY POLICE DEPARTMENT

LAB. DATA MASTER SERIAL NUMBER 091230
11/20/10


SUPPLY NUMBER: 0000

BLANK TEST	0.00%
INTERNAL STANDARD	0.00%
EXTERNAL STANDARD	0.00%
BLANK TEST	0.00%
EXTERNAL STANDARD	0.00%
BLANK TEST	0.00%
EXTERNAL STANDARD	0.00%
BLANK TEST	0.00%
EXTERNAL STANDARD	0.00%
BLANK TEST	0.00%
EXTERNAL STANDARD	0.00%
BLANK TEST	0.00%
EXTERNAL STANDARD	0.00%

PERCENTAGE: 0.00%

LABORATORY POLICE DEPARTMENT
MISSISSIPPI DEPARTMENT OF TRANSPORTATION
LABORATORY

Operator Signature



2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI
LABORATORY DIVISION

MISSISSIPPI STATE POLICE - MEMPHIS
LABORATORY

PROJECT NUMBER: 1000000000
SUBJECT: MURDER
CASE NUMBER: 1000000000
STATE POLICE - MEMPHIS
LABORATORY

DATE RECEIVED: 10/10/00
RECEIVED BY: [Signature]
ANALYST: [Signature]
LABORATORY: MEMPHIS
EVIDENCE NUMBER: 1000000000

LABORATORY: MEMPHIS

RECEIVED BY: [Signature] DATE: 10/10/00
LABORATORY: MEMPHIS STATE POLICE - MEMPHIS
EVIDENCE NUMBER: 1000000000

Operator Signature 