



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED** REPORT #6  
 By Carol Day at 2:14 pm, Jan 06, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201239	NAME OF AGENCY Lebanon Police Department	DATE OF INSPECTION 12/29/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 401 S Jefferson, Lebanon, Mo. 65536		TIME OF INSPECTION 8:02 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12/29/2014 @ 08:02 hrs
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing LOT # 13002 EXP. DATE 06/19/2015
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2755 EXP. DATE 08/12/2015

CALIBRATION CHECK -- (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .096	TEST 2 .097	TEST 3 .098
-------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	2	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is working within DOH specifications.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME Matthew S. Wright
TYPE II PERMIT NUMBER/EXPIRATION DATE 230033 02/19/2015	TELEPHONE NUMBER (417) 532-3131

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 13002**

**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



## Simulator Calibration Report

This calibration report is to certify the alcohol reference simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance with the standards set by the Missouri Department of Health and Senior Services Rules and Regulations: 19 CSR 25-30.051 (4).

Checked: 08/12/2014 Expires: 08/12/2015  
Digital Therm. SN:094948  
MSC Tech:DRL Temp:34.00  
Agency: Lebanon Police Dept  
SD 2755



Technician Printed Name: DAW LUCAS

Technician Signature: *DRL*

Date: 8/12/2014

Contact: Missouri Safety Center

Breath-Alcohol Instrument Training Program

660-543-4834



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**MATTHEW S WRIGHT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 02/19/2013

NUMBER 230033

EXPIRES 02/19/2015

*MS Wright*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Acting Director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)

# BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI

DEPARTMENT OF TRANSPORTATION

NEW ORLEANS DISTRICT OFFICE, FORTBERRY BUILDING  
TULSA, LA

VEHICLE IDENTIFICATION  
NUMBER: 1234567890  
DATE OF TEST: 10/15/12  
TESTER: J. SMITH  
LABORATORY: MISSISSIPPI  
DEPARTMENT OF TRANSPORTATION

--- ALCOHOL CONCENTRATION ---

DRIVER - 1234	0.08%	0.08%
ADJUTANT - 567890	0.02%	0.02%
ENROUTE - 1234567	0.00%	0.00%
BIOPHASE - 8901234	0.00%	0.00%
ENROUTE - 5678901	0.00%	0.00%
BIOPHASE - 2345678	0.00%	0.00%
ENROUTE - 9012345	0.00%	0.00%
BIOPHASE - 6789012	0.00%	0.00%
ENROUTE - 3456789	0.00%	0.00%
BIOPHASE - 0123456	0.00%	0.00%

TESTER  
SUPERVISOR  
LABORATORY

Operator Signature *[Signature]* 2208-02

# BAC DataMaster Evidence Ticket

STATE OF MISSISSIPPI

DEPARTMENT OF TRANSPORTATION

NEW ORLEANS DISTRICT OFFICE, FORTBERRY BUILDING  
TULSA, LA

--- ALCOHOL CONCENTRATION ---

DRIVER - 1234	0.08%	0.08%
ADJUTANT - 567890	0.02%	0.02%
ENROUTE - 1234567	0.00%	0.00%
BIOPHASE - 8901234	0.00%	0.00%
ENROUTE - 5678901	0.00%	0.00%
BIOPHASE - 2345678	0.00%	0.00%
ENROUTE - 9012345	0.00%	0.00%
BIOPHASE - 6789012	0.00%	0.00%
ENROUTE - 3456789	0.00%	0.00%
BIOPHASE - 0123456	0.00%	0.00%

PRELIMINARY RESULTS  
The above information was obtained from a preliminary test of the driver's blood sample. The results are subject to change upon receipt of the final laboratory report.

Operator Signature *[Signature]* 2208-02

