



Complete this report at the time of the regular monthly preventive maintenance check (not to be completed at the time of a repair or service).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed back into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

**REVIEWED**  
 By Carol Day at 1:36 pm, Feb 04, 2014

DATAMASTER SN 201241	NAME OF AGENCY TOWN AND COUNTRY POLICE DEPARTMENT	DATE OF INSPECTION 01/21/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 1011 MUNICIPAL CENTER DRIVE TOWN & COUNTRY, MO 63131		TIME OF INSPECTION 7:17 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>01/21/2014 19:17</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50 °C</u>	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>REPCO MARKETING INC</u> LOT # <u>13002</u> EXP. DATE <u>06/19/2015</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP ( $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ ) <u>34.00</u> °C SIMULATOR SN <u>SD2781</u> EXP. DATE <u>11/21/2014</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within  $\pm 5\%$  of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 $\leftarrow$ .099%	TEST 2 $\leftarrow$ .098%	TEST 3 $\leftarrow$ .097%
---------------------------	---------------------------	---------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	5	(.0-.04)	0	(.05-.09)	3	(.10-.14)	5	(.15-.19)	7	OVER .19	2
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME WM. CHRIS MOORE
TYPE II PERMIT NUMBER/EXPIRATION DATE 220302 09/25/2014	TELEPHONE NUMBER (314) 432-4697

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**REPCO MARKETING INC.**

3101-198 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-8480

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**  
**LOT NUMBER: 13002**  
**EXPIRATION DATE: June 19, 2015 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
TOWN AND COUNTRY POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 201241  
01/21/14

ARREST TIME: 19:08  
SUBJECT NAME: MOUSE/MICKY  
DOB: 11/11/11 SEX: M  
STATE/D.L.: FL/1E34567  
ARRESTING OFFICER: MOORE/WA/CHRIS  
OFFICER I.D.: 83  
TESTING OFFICER: MOORE/WA/CHRIS  
OFFICER I.D.: 83  
PERMIT NUMBER: 228202  
EXPIRATION DATE: 09/25/14  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	19:27
INTERNAL STANDARD	VERIFIED	19:27
RADIO INTERFERENCE		

Operator Signature

State of Missouri  
DEPARTMENT OF HEALTH



P E R M I T  
TYPE II



WILLIAM C MOORE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/25/2012

Number 220302

Expires 09/25/2014

Director of State Public Health Laboratory

Director, Department of Health