



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

REPORT #6

By Carol Day at 11:57 am, Apr 07, 2014

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201246	NAME OF AGENCY Monett PD	DATE OF INSPECTION 04/07/2014
LOCATION OF INSTRUMENT (STREET AND CITY) 1901 East Cleveland Monett		TIME OF INSPECTION 2:45 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 02:46
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth LOT # 13290 EXP. DATE 10/29/2015

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2267 EXP. DATE 07/10/2014

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← 0.101	TEST 2 ← 0.102	TEST 3 ← 0.101
----------------	----------------	----------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	3	(0-.04)	0	(.05-.09)	3	(.10-.14)	3	(.15-.19)	3	OVER .19	3
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Performed to DHSS Specs. All PBT's Calibrated (3 FST)

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME Ryan Hutton
TYPE II PERMITS NUMBER/EXPIRATION DATE 220425 12/27/2014	TELEPHONE NUMBER (417) 235-4241

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
MONETT POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201246
01 07 14

ARREST TIME: 08100
SUBJECT NAME:
MOUSE, NICKY
DOB: 12-12-90 SEX: M
STATE/D.L.# NAME
ARRESTING OFFICER:
HUTTON, RYAN L
OFFICER I.D.# 403
TESTING OFFICER:
HUTTON, RYAN L
OFFICER I.D.# 403
PERMIT NUMBER: 220425
EXPIRATION DATE: 12/31/14
MISCELLANEOUS DATA:
RFI TEST

BREATH ANALYZER

BLANK TEST 0000 0000
INTERNAL STANDARD VERIFIED 0000
RADIO INTERFERENCE

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

STATE OF MISSOURI
MONETT POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 281046
04/27/14

TESTING OFFICER:
HUTTON, RYAN L
OFFICER T.D. # 323
PERMIT NUMBER: 220403
EXPIRATION DATE: 12-27-14
MISCELLANEOUS DATA:
84734 RD

--- SUPERVISOR MODE ---

BLANK TEST	.000	02156
INTERNAL STANDARD	VERIFIED	02156
EXTERNAL STANDARD	.101	02156
BLANK TEST	.000	02157
EXTERNAL STANDARD	.100	02157
BLANK TEST	.000	02158
EXTERNAL STANDARD	.101	02158
BLANK TEST	.000	02159

N = 3
SD = .1
RVD = 11013

OPERATOR SIGNATURE

R. Hutton 323

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1495, MANITOWISH, WI 53048